

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103024\
 Data File : VN084592.D
 Acq On : 31 Oct 2024 01:45
 Operator : JC\MD
 Sample : P4597-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BLUE-2-2

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/04/2024
 Supervised By :Mahesh Dadoda 11/04/2024

Quant Time: Oct 31 18:32:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:24:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	188773	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	346573	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	312102	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	135645	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	136900	50.211	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 100.420%			
35) Dibromofluoromethane	8.165	113	111094	47.357	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 94.720%			
50) Toluene-d8	10.565	98	416629	48.213	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 96.420%			
62) 4-Bromofluorobenzene	12.847	95	150883	46.719	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 93.440%			
Target Compounds						
16) Acetone	4.442	43	19509	22.171	ug/l	90
20) Methylene Chloride	5.265	84	2409m	1.005	ug/l	
43) Isopropyl Acetate	8.688	43	155689	28.127	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103024\
Data File : VN084592.D
Acq On : 31 Oct 2024 01:45
Operator : JC\MD
Sample : P4597-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BLUE-2-2

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/04/2024
Supervised By :Mahesh Dadoda 11/04/2024

Quant Time: Oct 31 18:32:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:24:48 2024
Response via : Initial Calibration

