

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103024\
 Data File : VN084593.D
 Acq On : 31 Oct 2024 02:09
 Operator : JC\MD
 Sample : P4598-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-F12

Quant Time: Nov 01 05:18:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	184609	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	341082	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	315281	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	126962	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	137480	51.561	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	103.120%	
35) Dibromofluoromethane	8.165	113	111163	48.149	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	96.300%	
50) Toluene-d8	10.565	98	407989	47.973	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	95.940%	
62) 4-Bromofluorobenzene	12.847	95	148981	46.873	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	93.740%	
Target Compounds						
16) Acetone	4.424	43	17326	19.869	ug/l	Qvalue 94
43) Isopropyl Acetate	8.688	43	134942	24.654	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103024\
Data File : VN084593.D
Acq On : 31 Oct 2024 02:09
Operator : JC\MD
Sample : P4598-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-F12

Quant Time: Nov 01 05:18:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

