

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103024\
 Data File : VN084605.D
 Acq On : 31 Oct 2024 06:57
 Operator : JC\MD
 Sample : PB164501ZHE#14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB164501ZHE#14

Quant Time: Nov 01 06:18:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Semsettin
 Yesilyurt

11/04/2024
 Supervised By :Mahesh
 Dadoda

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	171568	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	311127	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	285569	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	121995	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	127842	51.590	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.180%
35) Dibromofluoromethane	8.165	113	100726	47.829	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.660%
50) Toluene-d8	10.565	98	373263	48.115	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.240%
62) 4-Bromofluorobenzene	12.847	95	132325	45.640	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	91.280%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	25254	32.802	ug/l	93
18) Methyl Acetate	5.036	43	15097	1.937	ug/l	98
20) Methylene Chloride	5.265	84	2566	1.177	ug/l #	67
43) Isopropyl Acetate	8.688	43	201204m	40.924	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103024\
Data File : VN084605.D
Acq On : 31 Oct 2024 06:57
Operator : JC\MD
Sample : PB164501ZHE#14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB164501ZHE#14

Manual Integrations
APPROVED

Reviewed By :Semsettin
Yesilyurt

11/04/2024
Supervised By :Mahesh
Dadoda

11/04/2024

Quant Time: Nov 01 06:18:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

