

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103024\
 Data File : VN084607.D
 Acq On : 31 Oct 2024 07:46
 Operator : JC\MD
 Sample : PB164501ZHE#16
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB164501ZHE#16

Quant Time: Nov 01 06:19:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	173914	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	322125	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	291140	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	123523	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	129490	51.551	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.100%			
35) Dibromofluoromethane	8.165	113	102414	46.970	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 93.940%			
50) Toluene-d8	10.565	98	377424	46.991	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 93.980%			
62) 4-Bromofluorobenzene	12.847	95	133845	44.589	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 89.180%			
Target Compounds						
16) Acetone	4.436	43	26913	34.634	ug/l	96
18) Methyl Acetate	5.018	43	14002	1.403	ug/l #	67
20) Methylene Chloride	5.271	84	2574	1.165	ug/l #	77
43) Isopropyl Acetate	8.688	43	222325m	43.742	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103024\
Data File : VN084607.D
Acq On : 31 Oct 2024 07:46
Operator : JC\MD
Sample : PB164501ZHE#16
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB164501ZHE#16

Quant Time: Nov 01 06:19:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

