

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103024\
 Data File : VN084609.D
 Acq On : 31 Oct 2024 08:34
 Operator : JC\MD
 Sample : PB164501ZHE#18
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB164501ZHE#18

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/04/2024
 Supervised By :Mahesh Dadoda 11/04/2024

Quant Time: Nov 01 06:20:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	170711	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	312595	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	289877	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	123376	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	127845	51.851	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	103.700%		
35) Dibromofluoromethane	8.159	113	101167	47.813	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.620%		
50) Toluene-d8	10.565	98	380435	48.810	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	97.620%		
62) 4-Bromofluorobenzene	12.847	95	130767	44.891	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	89.780%		
Target Compounds						
16) Acetone	4.436	43	27534	36.220	ug/l	99
18) Methyl Acetate	5.036	43	16067	2.373	ug/l #	83
20) Methylene Chloride	5.265	84	2331	1.075	ug/l #	65
43) Isopropyl Acetate	8.688	43	200107m	40.500	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103024\
Data File : VN084609.D
Acq On : 31 Oct 2024 08:34
Operator : JC\MD
Sample : PB164501ZHE#18
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB164501ZHE#18

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/04/2024
Supervised By :Mahesh Dadoda 11/04/2024

Quant Time: Nov 01 06:20:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

