

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103024\
 Data File : VN084611.D
 Acq On : 31 Oct 2024 09:22
 Operator : JC\MD
 Sample : PB164501ZHE#20
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB164501ZHE#20

Quant Time: Nov 01 06:21:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	163480	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	305291	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	275900	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	118079	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	122513	51.886	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.780%			
35) Dibromofluoromethane	8.165	113	98404	47.620	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 95.240%			
50) Toluene-d8	10.565	98	371443	48.796	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 97.600%			
62) 4-Bromofluorobenzene	12.847	95	127790	44.919	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 89.840%			
Target Compounds						
16) Acetone	4.436	43	31557	43.916	ug/l	89
18) Methyl Acetate	5.024	43	17356	3.232	ug/l	98
20) Methylene Chloride	5.265	84	2636	1.269	ug/l #	82
43) Isopropyl Acetate	8.688	43	209935m	43.579	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103024\
Data File : VN084611.D
Acq On : 31 Oct 2024 09:22
Operator : JC\MD
Sample : PB164501ZHE#20
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB164501ZHE#20

Quant Time: Nov 01 06:21:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

