

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103124\
 Data File : VN084623.D
 Acq On : 31 Oct 2024 17:01
 Operator : JC\MD
 Sample : P4646-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 241030043-05-TRIP-BLANK

Quant Time: Nov 01 02:49:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N103124W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Nov 01 02:38:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.806	128	30549	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.100	114	160545	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.865	117	142191	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.577	65	72870	29.671	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	98.900%
60) 4-Bromofluorobenzene	12.847	95	61359	25.062	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	83.533%
63) Toluene-d8	10.565	98	188693	28.032	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	93.433%
Target Compounds						
12) Methyl Acetate	5.036	43	4714	1.699	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103124\
Data File : VN084623.D
Acq On : 31 Oct 2024 17:01
Operator : JC\MD
Sample : P4646-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
241030043-05-TRIP-BLANK

Quant Time: Nov 01 02:49:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N103124W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Nov 01 02:38:04 2024
Response via : Initial Calibration

