

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110124\
 Data File : VN084629.D
 Acq On : 01 Nov 2024 11:04
 Operator : JC\MD
 Sample : VN1101MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1101MBL01

Quant Time: Nov 04 04:23:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	152334	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	265565	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	232363	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	100670	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	114355	51.974	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.940%			
35) Dibromofluoromethane	8.165	113	91394	50.844	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.680%			
50) Toluene-d8	10.565	98	315692	47.676	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 95.360%			
62) 4-Bromofluorobenzene	12.847	95	111797	45.176	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 90.360%			
Target Compounds						
					Qvalue	
77) Bromobenzene	12.988	156	2014	1.062	ug/l	68
82) 4-Chlorotoluene	13.223	91	5700	1.003	ug/l	98
87) 1,3-Dichlorobenzene	13.729	146	5255	1.420	ug/l	90
88) 1,4-Dichlorobenzene	13.811	146	6291	1.010	ug/l	84
89) n-Butylbenzene	14.058	91	5793	1.105	ug/l	97
93) 1,2,4-Trichlorobenzene	15.394	180	3969	2.039	ug/l	88
94) Hexachlorobutadiene	15.494	225	1226	1.434	ug/l	92
95) Naphthalene	15.635	128	7714	1.324	ug/l	97
96) 1,2,3-Trichlorobenzene	15.841	180	3358	1.777	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110124\
Data File : VN084629.D
Acq On : 01 Nov 2024 11:04
Operator : JC\MD
Sample : VN1101MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1101MBL01

Quant Time: Nov 04 04:23:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

