

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110124\
Data File : VN084635.D
Acq On : 01 Nov 2024 13:53
Operator : JC\MD
Sample : P4665-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-190-TB-20241029

Quant Time: Nov 04 04:27:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.223	168	171981	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	304211	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.864	117	270682	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	117747	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	126157	50.788	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	101.580%	
35) Dibromofluoromethane	8.165	113	101331	49.210	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	98.420%	
50) Toluene-d8	10.565	98	356786	47.037	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	94.080%	
62) 4-Bromofluorobenzene	12.847	95	134831	47.562	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	95.120%	
Target Compounds						
29) Tetrahydrofuran	7.841	42	7157	9.309	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110124\
Data File : VN084635.D
Acq On : 01 Nov 2024 13:53
Operator : JC\MD
Sample : P4665-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-190-TB-20241029

Quant Time: Nov 04 04:27:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

