

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110424\
 Data File : VN084646.D
 Acq On : 04 Nov 2024 10:14
 Operator : JC\MD
 Sample : VN1104MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1104MBL01

Quant Time: Nov 05 00:20:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	157878	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	275013	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	235777	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	103376	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	112531	49.349	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	98.700%	
35) Dibromofluoromethane	8.159	113	94390	50.706	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	101.420%	
50) Toluene-d8	10.565	98	319192	46.549	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	93.100%	
62) 4-Bromofluorobenzene	12.847	95	114001	44.484	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	88.960%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110424\
Data File : VN084646.D
Acq On : 04 Nov 2024 10:14
Operator : JC\MD
Sample : VN1104MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1104MBL01

Quant Time: Nov 05 00:20:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

