

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110424\
 Data File : VN084654.D
 Acq On : 04 Nov 2024 13:53
 Operator : JC\MD
 Sample : P4643-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-9

Quant Time: Nov 05 00:25:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	171283	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	310101	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	279437	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	116775	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	127838	51.675	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.340%
35) Dibromofluoromethane	8.165	113	102153	48.667	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.340%
50) Toluene-d8	10.565	98	375610	48.578	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.160%
62) 4-Bromofluorobenzene	12.847	95	131655	45.560	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	91.120%
Target Compounds						
16) Acetone	4.430	43	18432	23.205	ug/l #	89
43) Isopropyl Acetate	8.688	43	119770	24.045	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110424\
Data File : VN084654.D
Acq On : 04 Nov 2024 13:53
Operator : JC\MD
Sample : P4643-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-9

Quant Time: Nov 05 00:25:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

