

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110424\
 Data File : VN084661.D
 Acq On : 04 Nov 2024 16:41
 Operator : JC\MD
 Sample : P4660-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-CONCRETE-01-G

Quant Time: Nov 05 00:30:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	183096	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	318203	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	292659	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	141822	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	122924	46.482	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	92.960%		
35) Dibromofluoromethane	8.159	113	99331	46.118	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.240%		
50) Toluene-d8	10.565	98	377748	47.611	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	95.220%		
62) 4-Bromofluorobenzene	12.847	95	152813	51.535	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	103.060%		
Target Compounds						
16) Acetone	4.436	43	67807	86.903	ug/l	98
43) Isopropyl Acetate	8.688	43	132447	25.989	ug/l #	83
95) Naphthalene	15.635	128	629152	76.644	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110424\
Data File : VN084661.D
Acq On : 04 Nov 2024 16:41
Operator : JC\MD
Sample : P4660-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-CONCRETE-01-G

Quant Time: Nov 05 00:30:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

