

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110424\
 Data File : VN084667.D
 Acq On : 04 Nov 2024 19:06
 Operator : JC\MD
 Sample : P4679-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-1

Quant Time: Nov 05 00:33:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	180567	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	318837	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	289897	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	130368	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	125085	47.962	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	95.920%	
35) Dibromofluoromethane	8.165	113	101439	47.003	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	94.000%	
50) Toluene-d8	10.565	98	380650	47.881	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	95.760%	
62) 4-Bromofluorobenzene	12.847	95	143664	48.353	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	96.700%	
Target Compounds						
16) Acetone	4.436	43	59260	76.684	ug/l	95
43) Isopropyl Acetate	8.682	43	155202	30.561	ug/l #	82
95) Naphthalene	15.641	128	23557	3.122	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110424\
Data File : VN084667.D
Acq On : 04 Nov 2024 19:06
Operator : JC\MD
Sample : P4679-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-1

Quant Time: Nov 05 00:33:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

