

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110624\
 Data File : VN084703.D
 Acq On : 06 Nov 2024 11:48
 Operator : JC\MD
 Sample : VN1106MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1106MBL01

Quant Time: Nov 07 00:15:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	169887	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	295328	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	251754	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	109140	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	119604	48.744	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	97.480%	
35) Dibromofluoromethane	8.165	113	97421	48.735	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	97.460%	
50) Toluene-d8	10.565	98	339723	46.135	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	92.260%	
62) 4-Bromofluorobenzene	12.847	95	123560	44.897	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	89.800%	

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110624\
Data File : VN084703.D
Acq On : 06 Nov 2024 11:48
Operator : JC\MD
Sample : VN1106MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1106MBL01

Quant Time: Nov 07 00:15:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

