

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110624\
Data File : VN084710.D
Acq On : 06 Nov 2024 14:50
Operator : JC\MD
Sample : P4710-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-BPOW6-11-GW-20241031

Quant Time: Nov 07 00:19:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	178323	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	310385	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	273551	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	118494	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	123275	47.863	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	95.720%	
35) Dibromofluoromethane	8.159	113	102809	48.935	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	97.860%	
50) Toluene-d8	10.565	98	360006	46.517	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	93.040%	
62) 4-Bromofluorobenzene	12.853	95	137261	47.456	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	94.920%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110624\
Data File : VN084710.D
Acq On : 06 Nov 2024 14:50
Operator : JC\MD
Sample : P4710-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-BPOW6-11-GW-20241031

Quant Time: Nov 07 00:19:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

