

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110624\
 Data File : VN084712.D
 Acq On : 06 Nov 2024 15:38
 Operator : JC\MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VIBLK

Quant Time: Nov 07 00:20:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	184210	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	322841	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	281664	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	129671	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	128373	48.249	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	96.500%	
35) Dibromofluoromethane	8.165	113	104438	47.792	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	95.580%	
50) Toluene-d8	10.565	98	369758	45.934	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	91.860%	
62) 4-Bromofluorobenzene	12.847	95	140117	46.574	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	93.140%	

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110624\
Data File : VN084712.D
Acq On : 06 Nov 2024 15:38
Operator : JC\MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VIBLK

Quant Time: Nov 07 00:20:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

