

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110724\
 Data File : VN084721.D
 Acq On : 07 Nov 2024 14:58
 Operator : JC\MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VIBLK

Quant Time: Nov 08 00:26:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	154885	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	270130	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	229489	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	96785	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	109584	48.986	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	97.980%	
35) Dibromofluoromethane	8.165	113	91559	50.075	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	100.140%	
50) Toluene-d8	10.565	98	309456	45.944	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	91.880%	
62) 4-Bromofluorobenzene	12.847	95	111988	44.488	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	88.980%	

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110724\
Data File : VN084721.D
Acq On : 07 Nov 2024 14:58
Operator : JC\MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VIBLK

Quant Time: Nov 08 00:26:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

