

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110724\
 Data File : VN084733.D
 Acq On : 07 Nov 2024 19:58
 Operator : JC\MD
 Sample : P4759-01 20X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MONTCLAIR-TOTE-1

Quant Time: Nov 08 00:35:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	181938	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	329677	50.000	ug/l	# 0.00
63) Chlorobenzene-d5	11.865	117	287441	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	139551	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	124266	47.289	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.580%
35) Dibromofluoromethane	8.159	113	103570	46.412	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.820%
50) Toluene-d8	10.565	98	366823	44.625	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.240%
62) 4-Bromofluorobenzene	12.847	95	166622	54.236	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	108.480%
Target Compounds						
16) Acetone	4.430	43	113360	148.179	ug/l	99
25) 2-Butanone	7.489	43	17464	14.245	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110724\
Data File : VN084733.D
Acq On : 07 Nov 2024 19:58
Operator : JC\MD
Sample : P4759-01 20X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MONTCLAIR-TOTE-1

Quant Time: Nov 08 00:35:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

