

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120622\
 Data File : VN075578.D
 Acq On : 06 Dec 2022 12:13
 Operator : JC\MD
 Sample : VN1206MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1206MBL01

Quant Time: Dec 07 00:11:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	129964	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	217572	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	188987	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	78499	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	38003	54.475	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.940%
35) Dibromofluoromethane	8.171	113	46982	54.969	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.940%
50) Toluene-d8	10.571	98	93681	48.665	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.340%
62) 4-Bromofluorobenzene	12.853	95	63138	44.436	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120622\
Data File : VN075578.D
Acq On : 06 Dec 2022 12:13
Operator : JC\MD
Sample : VN1206MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1206MBL01

Quant Time: Dec 07 00:11:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

