

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120622\
 Data File : VN075582.D
 Acq On : 06 Dec 2022 13:48
 Operator : JC\MD
 Sample : N5868-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-0

Quant Time: Dec 07 00:13:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	189564	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	318644	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	287993	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	124815	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	52492	51.587	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 103.180%		
35) Dibromofluoromethane	8.171	113	64093	51.203	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.400%		
50) Toluene-d8	10.570	98	142225	50.448	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 100.900%		
62) 4-Bromofluorobenzene	12.847	95	101897	48.967	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 97.940%		
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	12775	13.460	ug/l	95
20) Methylene Chloride	5.294	84	5584	1.665	ug/l	88
37) Ethyl Acetate	7.571	43	21679	8.820	ug/l	99
43) Isopropyl Acetate	8.694	43	109077	27.647	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120622\
Data File : VN075582.D
Acq On : 06 Dec 2022 13:48
Operator : JC\MD
Sample : N5868-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-0

Quant Time: Dec 07 00:13:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

