

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120622\
 Data File : VN075588.D
 Acq On : 06 Dec 2022 16:38
 Operator : JC\MD
 Sample : N5876-02DL 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 LIN-VRG-12-02DL

Quant Time: Dec 07 00:16:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	215914	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	341724	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	308851	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	145501	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	57648	49.740	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 99.480%			
35) Dibromofluoromethane	8.171	113	74622	55.588	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 111.180%			
50) Toluene-d8	10.571	98	159292	52.685	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 105.380%			
62) 4-Bromofluorobenzene	12.853	95	118932	53.293	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 106.580%			
Target Compounds						Qvalue
8) Diethyl Ether	3.971	74	4056	3.129	ug/l	87
16) Acetone	4.436	43	194867	180.256	ug/l	97
18) Methyl Acetate	5.036	43	16066	4.722	ug/l	100
25) 2-Butanone	7.488	43	966696	640.495	ug/l	100
29) Tetrahydrofuran	7.847	42	138363	150.713	ug/l	97
40) Benzene	8.612	78	226995	25.326	ug/l	98
52) Toluene	10.635	92	669406	115.897	ug/l	100
67) Ethyl Benzene	11.965	91	176466	16.217	ug/l	98
68) m/p-Xylenes	12.070	106	288889	65.375	ug/l	99
69) o-Xylene	12.400	106	134446	30.622	ug/l	99
73) Isopropylbenzene	12.700	105	15399	1.449	ug/l	96
78) n-propylbenzene	13.041	91	15149	1.277	ug/l	95
80) 1,3,5-Trimethylbenzene	13.176	105	22560	2.554	ug/l	99
84) 1,2,4-Trimethylbenzene	13.488	105	48121	5.432	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120622\
Data File : VN075588.D
Acq On : 06 Dec 2022 16:38
Operator : JC\MD
Sample : N5876-02DL 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
LIN-VRG-12-02DL

Quant Time: Dec 07 00:16:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

