

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121924\
 Data File : VN085268.D
 Acq On : 19 Dec 2024 16:23
 Operator : JC\MD
 Sample : P5328-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 001-WILLETS-PT-BLVD(DEC)

Quant Time: Dec 19 17:13:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N120624W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Dec 06 22:50:42 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.812	128	26278	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.100	114	117507	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.865	117	106486	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.576	65	72524	34.783	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	115.933%#
60) 4-Bromofluorobenzene	12.847	95	43274	24.561	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	81.867%
63) Toluene-d8	10.565	98	147845	29.598	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	98.667%
Target Compounds						
15) Acetone	4.424	58	56978	268.209	ug/l	83
30) 2-Butanone	7.482	43	11530	13.251	ug/l #	76
62) Toluene	10.629	91	352042	56.687	ug/l	97
82) p-Isopropyltoluene	13.723	119	2042	0.401	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121924\
Data File : VN085268.D
Acq On : 19 Dec 2024 16:23
Operator : JC\MD
Sample : P5328-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
001-WILLETS-PT-BLVD(DEC)

Quant Time: Dec 19 17:13:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N120624W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Dec 06 22:50:42 2024
Response via : Initial Calibration

