

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122721\
 Data File : VN070321.D
 Acq On : 27 Dec 2021 20:17
 Operator : JC/MD
 Sample : M5092-06 10X
 Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PATAPSCO-FEED-SLUDGE

Quant Time: Dec 28 01:44:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1214471	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	2341159	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2905679	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1179813	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	925725	53.01	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 106.02%			
35) Dibromofluoromethane	8.021	113	677294	47.81	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 95.62%			
50) Toluene-d8	10.440	98	3151487	54.46	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 108.92%			
62) 4-Bromofluorobenzene	12.731	95	1437238	68.77	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 137.54%#			
Target Compounds						
16) Acetone	4.296	43	9468	0.84	ug/l	92
52) Toluene	10.505	92	106421	2.45	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122721\
Data File : VN070321.D
Acq On : 27 Dec 2021 20:17
Operator : JC/MD
Sample : M5092-06 10X
Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PATAPSCO-FEED-SLUDGE

Quant Time: Dec 28 01:44:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

