

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122922\
 Data File : VN076044.D
 Acq On : 29 Dec 2022 13:19
 Operator : JC\MD
 Sample : PB149968TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149968TB

Quant Time: Dec 30 05:32:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	352398	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	550755	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	488978	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	201351	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	172409	44.971	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	89.940%		
35) Dibromofluoromethane	8.177	113	161056	47.220	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.440%		
50) Toluene-d8	10.571	98	642344	50.370	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	100.740%		
62) 4-Bromofluorobenzene	12.847	95	198878	47.086	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	94.180%		
Target Compounds					Qvalue	
16) Acetone	4.448	43	65555	49.604	ug/l	98
20) Methylene Chloride	5.289	84	20670	5.274	ug/l	98
25) 2-Butanone	7.494	43	24229	13.094	ug/l	95
37) Ethyl Acetate	7.571	43	34413	9.456	ug/l	100
43) Isopropyl Acetate	8.694	43	189060	31.492	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122922\
Data File : VN076044.D
Acq On : 29 Dec 2022 13:19
Operator : JC\MD
Sample : PB149968TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149968TB

Quant Time: Dec 30 05:32:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

