

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122922\
 Data File : VN076045.D
 Acq On : 29 Dec 2022 13:42
 Operator : JC\MD
 Sample : N6228-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1

Quant Time: Dec 30 05:32:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	372487	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	589938	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	530978	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	225990	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	186725	46.079	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.160%
35) Dibromofluoromethane	8.177	113	172360	47.178	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.360%
50) Toluene-d8	10.571	98	682165	49.940	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.880%
62) 4-Bromofluorobenzene	12.853	95	220180	48.667	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.340%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	43230	30.947	ug/l	97
20) Methylene Chloride	5.289	84	12498	2.739	ug/l	92
37) Ethyl Acetate	7.565	43	31941	8.194	ug/l	98
43) Isopropyl Acetate	8.694	43	186683	29.031	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122922\
Data File : VN076045.D
Acq On : 29 Dec 2022 13:42
Operator : JC\MD
Sample : N6228-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1

Quant Time: Dec 30 05:32:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

