

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122922\
 Data File : VN076050.D
 Acq On : 29 Dec 2022 15:42
 Operator : JC\MD
 Sample : N6231-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2S

Quant Time: Dec 30 05:34:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	369127	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	587612	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	525757	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	218669	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	183029	45.578	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.160%
35) Dibromofluoromethane	8.171	113	174268	47.889	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.780%
50) Toluene-d8	10.571	98	681891	50.118	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.240%
62) 4-Bromofluorobenzene	12.853	95	215529	47.827	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.660%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	42685	30.835	ug/l	93
20) Methylene Chloride	5.289	84	12513	2.774	ug/l	98
37) Ethyl Acetate	7.571	43	31703	8.165	ug/l	99
43) Isopropyl Acetate	8.694	43	263088	41.074	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122922\
Data File : VN076050.D
Acq On : 29 Dec 2022 15:42
Operator : JC\MD
Sample : N6231-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-2S

Quant Time: Dec 30 05:34:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

