

Data Path : Z:\voasrv\HPCHEM1\MSVOA_T\Data\VT102725\
 Data File : VT019186.D
 Acq On : 27 Oct 2025 9:16
 Operator : SY/MD
 Sample : VT1027WBL01
 Misc : 5.0mL/MSVOA_T/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_T
 ClientSampleId :
 VT1027WBL01

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 10/30/2025
 Supervised By :Mahesh Dadoda 10/30/2025

Quant Time: Oct 28 05:52:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_T\methods\624T102225W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Oct 24 06:05:07 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.031	128	23449m	30.000	ug/l	-0.05
28) 1,4-Difluorobenzene	8.535	114	152144	30.000	ug/l	-0.04
57) Chlorobenzene-d5	11.708	117	121465	30.000	ug/l	-0.05

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.918	65	63060	30.695	ug/l	-0.05
Spiked Amount	30.000	Range	91 - 110	Recovery	=	102.333%
60) 4-Bromofluorobenzene	12.842	95	54238	28.412	ug/l	-0.05
Spiked Amount	30.000	Range	63 - 112	Recovery	=	94.700%
63) Toluene-d8	10.209	98	179758	30.406	ug/l	-0.05
Spiked Amount	30.000	Range	91 - 112	Recovery	=	101.367%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_T\Data\VT102725\
Data File : VT019186.D
Acq On : 27 Oct 2025 9:16
Operator : SY/MD
Sample : VT1027WBL01
Misc : 5.0mL/MSVOA_T/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_T
ClientSampleId :
VT1027WBL01

Manual Integrations
APPROVED

Reviewed By :Amit Patel 10/30/2025
Supervised By :Mahesh Dadoda 10/30/2025

Quant Time: Oct 28 05:52:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_T\methods\624T102225W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Oct 24 06:05:07 2025
Response via : Initial Calibration

