

Data Path : Z:\voasrv\HPCHEM1\MSVOA_T\Data\VT102725\
 Data File : VT019187.D
 Acq On : 27 Oct 2025 9:39
 Operator : SY/MD
 Sample : VT1027WBL02
 Misc : 5.0mL/MSVOA_T/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_T
 ClientSampleId :
 VT1027WBL02

Quant Time: Oct 28 05:53:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_T\methods\624T102225W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Oct 24 06:05:07 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.030	128	25840	30.000	ug/l	#-0.05
28) 1,4-Difluorobenzene	8.529	114	174941	30.000	ug/l	-0.05
57) Chlorobenzene-d5	11.707	117	143098	30.000	ug/l	-0.05

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.918	65	70568	31.172	ug/l	-0.05
Spiked Amount	30.000	Range	91 - 110	Recovery	=	103.900%
60) 4-Bromofluorobenzene	12.835	95	61875	27.512	ug/l	-0.05
Spiked Amount	30.000	Range	63 - 112	Recovery	=	91.700%
63) Toluene-d8	10.209	98	203568	29.227	ug/l	-0.05
Spiked Amount	30.000	Range	91 - 112	Recovery	=	97.433%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_T\Data\VT102725\
Data File : VT019187.D
Acq On : 27 Oct 2025 9:39
Operator : SY/MD
Sample : VT1027WBL02
Misc : 5.0mL/MSVOA_T/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_T
ClientSampleId :
VT1027WBL02

Quant Time: Oct 28 05:53:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_T\methods\624T102225W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Oct 24 06:05:07 2025
Response via : Initial Calibration

