

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_T\Data\VT102925\  
 Data File : VT019199.D  
 Acq On : 29 Oct 2025 12:42  
 Operator : SY/MD  
 Sample : VSTDICV050  
 Misc : 5.0mL/MSVOA\_T/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_T  
 ClientSampleId :  
 ICVVT102925

Quant Time: Oct 30 08:45:28 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_T\methods\82T102925W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Oct 30 08:43:44 2025  
 Response via : Initial Calibration

### Manual Integrations APPROVED

Reviewed By :Amit Patel 10/30/2025  
 Supervised By :Mahesh Dadoda 10/30/2025

| Compound                     | R.T.   | QIon | Response | Conc    | Units    | Dev(Min) |
|------------------------------|--------|------|----------|---------|----------|----------|
| -----                        |        |      |          |         |          |          |
| Internal Standards           |        |      |          |         |          |          |
| 1) Pentafluorobenzene        | 7.554  | 168  | 143995   | 50.000  | ug/l     | 0.00     |
| 34) 1,4-Difluorobenzene      | 8.547  | 114  | 246224   | 50.000  | ug/l     | 0.00     |
| 63) Chlorobenzene-d5         | 11.725 | 117  | 199625   | 50.000  | ug/l     | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 13.929 | 152  | 98670    | 50.000  | ug/l     | 0.00     |
| System Monitoring Compounds  |        |      |          |         |          |          |
| 33) 1,2-Dichloroethane-d4    | 7.942  | 65   | 121609   | 52.505  | ug/l     | 0.00     |
| Spiked Amount 50.000         |        |      | Recovery | =       | 105.000% |          |
| 35) Dibromofluoromethane     | 7.471  | 113  | 81958    | 52.628  | ug/l     | 0.00     |
| Spiked Amount 50.000         |        |      | Recovery | =       | 105.260% |          |
| 50) Toluene-d8               | 10.233 | 98   | 306573   | 48.477  | ug/l     | 0.00     |
| Spiked Amount 50.000         |        |      | Recovery | =       | 96.960%  |          |
| 62) 4-Bromofluorobenzene     | 12.859 | 95   | 96220    | 48.398  | ug/l     | 0.00     |
| Spiked Amount 50.000         |        |      | Recovery | =       | 96.800%  |          |
| Target Compounds             |        |      |          |         |          | Qvalue   |
| 2) Dichlorodifluoromethane   | 1.866  | 85   | 42258    | 49.832  | ug/l     | 97       |
| 3) Chloromethane             | 2.107  | 50   | 60329m   | 46.161  | ug/l     |          |
| 4) Vinyl Chloride            | 2.183  | 62   | 73291    | 43.841  | ug/l     | 97       |
| 5) Bromomethane              | 2.512  | 94   | 30231m   | 33.571  | ug/l     |          |
| 6) Chloroethane              | 2.648  | 64   | 65775    | 43.364  | ug/l     | 95       |
| 7) Trichlorofluoromethane    | 2.959  | 101  | 132089   | 42.676  | ug/l     | 95       |
| 8) Diethyl Ether             | 3.335  | 74   | 76980    | 51.856  | ug/l     | 98       |
| 9) 1,1,2-Trichlorotrifluo... | 3.652  | 101  | 68808    | 45.820  | ug/l     | 93       |
| 10) Methyl Iodide            | 3.840  | 142  | 64222    | 52.823  | ug/l     | 97       |
| 11) Tert butyl alcohol       | 4.645  | 59   | 149086   | 302.826 | ug/l     | 99       |
| 12) 1,1-Dichloroethene       | 3.635  | 96   | 63149    | 45.513  | ug/l     | 99       |
| 13) Acrolein                 | 3.517  | 56   | 22348    | 266.096 | ug/l     | 95       |
| 14) Allyl chloride           | 4.193  | 41   | 134252   | 49.060  | ug/l     | 90       |
| 15) Acrylonitrile            | 4.816  | 53   | 333498   | 277.764 | ug/l     | 100      |
| 16) Acetone                  | 3.717  | 43   | 329410   | 238.380 | ug/l     | 99       |
| 17) Carbon Disulfide         | 3.934  | 76   | 86913    | 46.086  | ug/l     | 98       |
| 18) Methyl Acetate           | 4.199  | 43   | 233979   | 57.408  | ug/l     | 96       |
| 19) Methyl tert-butyl Ether  | 4.874  | 73   | 339660   | 54.196  | ug/l     | 97       |
| 20) Methylene Chloride       | 4.399  | 84   | 78253    | 50.535  | ug/l     | 97       |
| 21) trans-1,2-Dichloroethene | 4.869  | 96   | 56581    | 47.884  | ug/l     | 94       |
| 22) Diisopropyl ether        | 5.773  | 45   | 337565   | 51.962  | ug/l     | 97       |
| 23) Vinyl Acetate            | 5.715  | 43   | 1255223  | 270.453 | ug/l     | 99       |
| 24) 1,1-Dichloroethane       | 5.662  | 63   | 152175   | 50.073  | ug/l     | 99       |
| 25) 2-Butanone               | 6.667  | 43   | 456774   | 257.837 | ug/l     | 96       |
| 26) 2,2-Dichloropropane      | 6.655  | 77   | 118617   | 45.233  | ug/l     | 98       |
| 27) cis-1,2-Dichloroethene   | 6.661  | 96   | 87305    | 52.651  | ug/l     | 98       |
| 28) Bromochloromethane       | 7.048  | 49   | 87685    | 53.987  | ug/l #   | 99       |
| 29) Tetrahydrofuran          | 7.066  | 42   | 292657   | 283.927 | ug/l     | 91       |
| 30) Chloroform               | 7.236  | 83   | 146555   | 51.508  | ug/l     | 98       |
| 31) Cyclohexane              | 7.536  | 56   | 95937    | 42.453  | ug/l     | 94       |
| 32) 1,1,1-Trichloroethane    | 7.442  | 97   | 112387   | 49.622  | ug/l     | 99       |
| 36) 1,1-Dichloropropene      | 7.683  | 75   | 82130    | 46.887  | ug/l     | 99       |
| 37) Ethyl Acetate            | 6.772  | 43   | 167711   | 56.243  | ug/l     | 97       |
| 38) Carbon Tetrachloride     | 7.665  | 117  | 77966    | 48.952  | ug/l     | 99       |
| 39) Methylcyclohexane        | 9.093  | 83   | 88555    | 45.754  | ug/l     | 97       |
| 40) Benzene                  | 7.953  | 78   | 293881   | 49.562  | ug/l     | 97       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_T\Data\VT102925\  
 Data File : VT019199.D  
 Acq On : 29 Oct 2025 12:42  
 Operator : SY/MD  
 Sample : VSTDICV050  
 Misc : 5.0mL/MSVOA\_T/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_T  
 ClientSampleId :  
 ICVVT102925

Quant Time: Oct 30 08:45:28 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_T\methods\82T102925W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Oct 30 08:43:44 2025  
 Response via : Initial Calibration

### Manual Integrations APPROVED

Reviewed By :Amit Patel 10/30/2025  
 Supervised By :Mahesh Dadoda 10/30/2025

| Compound                      | R.T.   | QIon | Response | Conc     | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|----------|--------|----------|
| 41) Methacrylonitrile         | 7.019  | 41   | 76342    | 50.888   | ug/l   | 87       |
| 42) 1,2-Dichloroethane        | 8.041  | 62   | 111929   | 53.717   | ug/l   | 99       |
| 43) Isopropyl Acetate         | 8.094  | 43   | 259648   | 54.759   | ug/l # | 91       |
| 44) Trichloroethene           | 8.823  | 130  | 56194    | 48.482   | ug/l   | 96       |
| 45) 1,2-Dichloropropane       | 9.140  | 63   | 77519    | 51.381   | ug/l   | 97       |
| 46) Dibromomethane            | 9.240  | 93   | 48331    | 53.227   | ug/l   | 95       |
| 47) Bromodichloromethane      | 9.457  | 83   | 108495   | 52.999   | ug/l   | 98       |
| 48) Methyl methacrylate       | 9.228  | 41   | 115207   | 54.093   | ug/l   | 89       |
| 49) 1,4-Dioxane               | 9.234  | 88   | 44673    | 1116.760 | ug/l   | 95       |
| 51) 4-Methyl-2-Pentanone      | 10.110 | 43   | 790140   | 276.688  | ug/l   | 96       |
| 52) Toluene                   | 10.304 | 92   | 165395   | 48.815   | ug/l   | 97       |
| 53) t-1,3-Dichloropropene     | 10.556 | 75   | 116195   | 52.442   | ug/l   | 99       |
| 54) cis-1,3-Dichloropropene   | 9.945  | 75   | 120149   | 52.367   | ug/l   | 95       |
| 55) 1,1,2-Trichloroethane     | 10.762 | 97   | 71272    | 53.334   | ug/l   | 99       |
| 56) Ethyl methacrylate        | 10.615 | 69   | 140593   | 54.746   | ug/l   | 92       |
| 57) 1,3-Dichloropropane       | 10.926 | 76   | 125423   | 52.883   | ug/l   | 99       |
| 58) 2-Chloroethyl Vinyl ether | 9.786  | 63   | 185157   | 267.903  | ug/l   | 97       |
| 59) 2-Hexanone                | 10.979 | 43   | 564629   | 260.996  | ug/l   | 93       |
| 60) Dibromochloromethane      | 11.155 | 129  | 72737    | 53.986   | ug/l   | 100      |
| 61) 1,2-Dibromoethane         | 11.273 | 107  | 65140    | 52.349   | ug/l   | 98       |
| 64) Tetrachloroethene         | 10.844 | 164  | 42710    | 47.070   | ug/l   | 98       |
| 65) Chlorobenzene             | 11.755 | 112  | 176595   | 49.287   | ug/l   | 98       |
| 66) 1,1,1,2-Tetrachloroethane | 11.843 | 131  | 62079    | 50.606   | ug/l   | 98       |
| 67) Ethyl Benzene             | 11.843 | 91   | 313249   | 46.761   | ug/l   | 100      |
| 68) m/p-Xylenes               | 11.972 | 106  | 236279   | 95.489   | ug/l   | 100      |
| 69) o-Xylene                  | 12.342 | 106  | 118766   | 48.287   | ug/l   | 98       |
| 70) Styrene                   | 12.360 | 104  | 219989   | 49.670   | ug/l   | 98       |
| 71) Bromoform                 | 12.548 | 173  | 52382    | 55.844   | ug/l # | 98       |
| 73) Isopropylbenzene          | 12.683 | 105  | 315892   | 46.505   | ug/l   | 97       |
| 74) N-amyl acetate            | 12.472 | 43   | 181126   | 52.862   | ug/l   | 95       |
| 75) 1,1,2,2-Tetrachloroethane | 12.977 | 83   | 121280   | 52.108   | ug/l   | 96       |
| 76) 1,2,3-Trichloropropane    | 13.030 | 75   | 103936m  | 51.189   | ug/l   |          |
| 77) Bromobenzene              | 13.000 | 156  | 70543    | 48.946   | ug/l   | 92       |
| 78) n-propylbenzene           | 13.077 | 91   | 373758   | 46.314   | ug/l   | 97       |
| 79) 2-Chlorotoluene           | 13.171 | 91   | 218039   | 47.160   | ug/l   | 96       |
| 80) 1,3,5-Trimethylbenzene    | 13.230 | 105  | 255548   | 46.838   | ug/l   | 98       |
| 81) trans-1,4-Dichloro-2-b... | 12.742 | 75   | 48309    | 55.825   | ug/l   | 91       |
| 82) 4-Chlorotoluene           | 13.282 | 91   | 227000   | 48.249   | ug/l   | 98       |
| 83) tert-Butylbenzene         | 13.529 | 119  | 240426   | 47.610   | ug/l   | 96       |
| 84) 1,2,4-Trimethylbenzene    | 13.582 | 105  | 258900   | 46.724   | ug/l   | 97       |
| 85) sec-Butylbenzene          | 13.735 | 105  | 346167   | 46.734   | ug/l   | 95       |
| 86) p-Isopropyltoluene        | 13.864 | 119  | 279948   | 45.267   | ug/l   | 98       |
| 87) 1,3-Dichlorobenzene       | 13.864 | 146  | 140805   | 47.618   | ug/l   | 99       |
| 88) 1,4-Dichlorobenzene       | 13.952 | 146  | 142349   | 47.789   | ug/l   | 99       |
| 89) n-Butylbenzene            | 14.234 | 91   | 270954   | 47.433   | ug/l   | 97       |
| 90) Hexachloroethane          | 14.546 | 117  | 46022    | 48.236   | ug/l   | 93       |
| 91) 1,2-Dichlorobenzene       | 14.287 | 146  | 143274   | 49.459   | ug/l   | 98       |
| 92) 1,2-Dibromo-3-Chloropr... | 14.986 | 75   | 27557    | 54.900   | ug/l   | 94       |
| 93) 1,2,4-Trichlorobenzene    | 15.774 | 180  | 91896    | 49.324   | ug/l   | 99       |
| 94) Hexachlorobutadiene       | 15.897 | 225  | 36550    | 47.564   | ug/l   | 98       |
| 95) Naphthalene               | 16.062 | 128  | 338036   | 51.817   | ug/l   | 99       |
| 96) 1,2,3-Trichlorobenzene    | 16.297 | 180  | 91373    | 49.709   | ug/l   | 99       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_T\Data\VT102925\  
Data File : VT019199.D  
Acq On : 29 Oct 2025 12:42  
Operator : SY/MD  
Sample : VSTDICV050  
Misc : 5.0mL/MSVOA\_T/WATER  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVVT102925

Manual Integrations  
APPROVED

Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025

Quant Time: Oct 30 08:45:28 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_T\methods\82T102925W.M  
Quant Title : SW846 8260  
QLast Update : Thu Oct 30 08:43:44 2025  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

-----  
(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_T\Data\VT102925\  
Data File : VT019199.D  
Acq On : 29 Oct 2025 12:42  
Operator : SY/MD  
Sample : VSTDICV050  
Misc : 5.0mL/MSVOA\_T/WATER  
ALS Vial : 10 Sample Multiplier: 1


Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVVT102925

Manual Integrations  
APPROVED

Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025

Quant Time: Oct 30 08:45:28 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_T\methods\82T102925W.M  
Quant Title : SW846 8260  
QLast Update : Thu Oct 30 08:43:44 2025  
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.554 min Scan# 1029

Delta R.T. 0.006 min

Lab File: VT019199.D

Acq: 29 Oct 2025 12:42

Instrument :

MSVOA\_T

ClientSampleId :

ICVT102925

Tgt Ion:168 Resp: 14399

Ion Ratio Lower Upper

168 100





99 58.6 46.6 70.0

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/30/2025

Supervised By :Mahesh Dadoda 10/30/2025



#2

Dichlorodifluoromethane

Concen: 49.832 ug/l

RT: 1.866 min Scan# 63

Delta R.T. -0.006 min

Lab File: VT019199.D




Acq: 29 Oct 2025 12:42


Tgt Ion: 85 Resp: 42258

Ion Ratio Lower Upper

85 100

87 31.7 16.8 50.4









#3  
Chloromethane  
Concen: 46.161 ug/l m  
RT: 2.107 min Scan# 104  
Delta R.T. 0.041 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42

Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVT102925

Tgt Ion: 50 Resp: 60329  
Ion Ratio Lower Upper  
50 100  
52 29.5 26.5 39.7




Manual Integrations  
APPROVED


Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025



#4  
Vinyl Chloride  
Concen: 43.841 ug/l  
RT: 2.183 min Scan# 117  
Delta R.T. -0.006 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42


Tgt Ion: 62 Resp: 73291  
Ion Ratio Lower Upper  
62 100  
64 30.1 25.2 37.8





#5  
Bromomethane  
Concen: 33.571 ug/l m  
RT: 2.512 min Scan# 1  
Delta R.T. -0.006 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVT102925




Tgt Ion: 94 Resp: 3023  
Ion Ratio Lower Upper  
94 100  
96 92.3 70.8 106.2

Manual Integrations  
APPROVED



Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025




#6  
Chloroethane  
Concen: 43.364 ug/l  
RT: 2.648 min Scan# 196  
Delta R.T. -0.011 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Tgt Ion: 64 Resp: 65775  
Ion Ratio Lower Upper  
64 100  
66 30.6 26.9 40.3





#7  
 Trichlorofluoromethane  
 Concen: 42.676 ug/l  
 RT: 2.959 min Scan# 247  
 Delta R.T. -0.006 min  
 Lab File: VT019199.D  
 Acq: 29 Oct 2025 12:42




Instrument :  
 MSVOA\_T  
 ClientSampleId :  
 ICVVT102925



Tgt Ion:101 Resp: 132089  
 Ion Ratio Lower Upper  
 101 100  
 103 66.0 49.9 74.9




Manual Integrations  
 APPROVED

Reviewed By :Amit Patel 10/30/2025  
 Supervised By :Mahesh Dadoda 10/30/2025




#8  
 Diethyl Ether  
 Concen: 51.856 ug/l  
 RT: 3.335 min Scan# 313  
 Delta R.T. 0.000 min  
 Lab File: VT019199.D  
 Acq: 29 Oct 2025 12:42

Tgt Ion: 74 Resp: 76980  
 Ion Ratio Lower Upper  
 74 100  
 45 102.3 52.3 156.8











#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 45.820 ug/l  
RT: 3.652 min Scan# 30  
Delta R.T. 0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42

Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVVT102925

Tgt Ion:101 Resp: 6880  
Ion Ratio Lower Upper  
101 100  
85 45.2 37.0 55.6  
151 70.6 63.8 95.6




Manual Integrations  
APPROVED


Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025



#10  
Methyl Iodide  
Concen: 52.823 ug/l  
RT: 3.840 min Scan# 399  
Delta R.T. 0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42


Tgt Ion:142 Resp: 64222  
Ion Ratio Lower Upper  
142 100  
127 41.5 35.0 52.4  
141 13.8 11.4 17.2





#11  
Tert butyl alcohol  
Concen: 302.826 ug/l  
RT: 4.645 min Scan# 51  
Delta R.T. 0.006 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVT102925




Tgt Ion: 59 Resp: 149086  
Ion Ratio Lower Upper  
59 100  
57 9.8 8.1 12.1

Manual Integrations  
APPROVED



Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025




#12  
1,1-Dichloroethene  
Concen: 45.513 ug/l  
RT: 3.635 min Scan# 364  
Delta R.T. -0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Tgt Ion: 96 Resp: 63149  
Ion Ratio Lower Upper  
96 100  
61 178.2 141.4 212.2  
98 63.4 50.7 76.1





#13  
Acrolein  
Concen: 266.096 ug/l  
RT: 3.517 min Scan# 342  
Delta R.T. 0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVVT102925




Tgt Ion: 56 Resp: 22348  
Ion Ratio Lower Upper  
56 100  
55 70.2 59.4 89.0

Manual Integrations  
APPROVED



Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025



#14  
Allyl chloride  
Concen: 49.060 ug/l  
RT: 4.193 min Scan# 459  
Delta R.T. 0.006 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Tgt Ion: 41 Resp: 134252  
Ion Ratio Lower Upper  
41 100  
39 69.9 49.9 74.9  
76 31.1 20.8 31.2





#15  
Acrylonitrile  
Concen: 277.764 ug/l  
RT: 4.816 min Scan# 50  
Delta R.T. -0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVVT102925



Tgt Ion: 53 Resp: 333498  
Ion Ratio Lower Upper  
53 100  
52 82.7 65.9 98.9  
51 36.4 29.4 44.2




Manual Integrations  
APPROVED


Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025



#16  
Acetone  
Concen: 238.380 ug/l  
RT: 3.717 min Scan# 378  
Delta R.T. -0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42


Tgt Ion: 43 Resp: 329410  
Ion Ratio Lower Upper  
43 100  
58 30.4 23.8 35.8





#17  
Carbon Disulfide  
Concen: 46.086 ug/l  
RT: 3.934 min Scan# 41  
Delta R.T. 0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVT102925



Tgt Ion: 76 Resp: 8691  
Ion Ratio Lower Upper  
76 100  
78 9.8 7.1 10.7




Manual Integrations  
APPROVED


Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025



#18  
Methyl Acetate  
Concen: 57.408 ug/l  
RT: 4.199 min Scan# 460  
Delta R.T. 0.006 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42


Tgt Ion: 43 Resp: 233979  
Ion Ratio Lower Upper  
43 100  
74 22.9 16.7 25.1





#19  
Methyl tert-butyl Ether  
Concen: 54.196 ug/l  
RT: 4.874 min Scan# 51  
Delta R.T. 0.006 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVVT102925



Tgt Ion: 73 Resp: 339660  
Ion Ratio Lower Upper  
73 100  
57 21.8 18.8 28.2




Manual Integrations  
APPROVED


Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025



#20  
Methylene Chloride  
Concen: 50.535 ug/l  
RT: 4.399 min Scan# 494  
Delta R.T. 0.001 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42


Tgt Ion: 84 Resp: 78253  
Ion Ratio Lower Upper  
84 100  
49 137.2 113.1 169.7  
51 39.4 33.6 50.4  
86 62.3 50.6 75.8





#21  
trans-1,2-Dichloroethene  
Concen: 47.884 ug/l  
RT: 4.869 min Scan# 51  
Delta R.T. 0.007 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVVT102925




Tgt Ion: 96 Resp: 5658  
Ion Ratio Lower Upper  
96 100  
61 146.0 124.6 187.0  
98 65.9 50.9 76.3

Manual Integrations  
APPROVED



Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025




#22  
Diisopropyl ether  
Concen: 51.962 ug/l  
RT: 5.773 min Scan# 728  
Delta R.T. 0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42



Tgt Ion: 45 Resp: 337565  
Ion Ratio Lower Upper  
45 100  
43 69.3 54.2 81.4  
87 25.4 18.9 28.3  
59 12.8 8.7 13.1









#23  
 Vinyl Acetate  
 Concen: 270.453 ug/l  
 RT: 5.715 min Scan# 716  
 Delta R.T. 0.006 min  
 Lab File: VT019199.D  
 Acq: 29 Oct 2025 12:42

Instrument :  
 MSVOA\_T  
 ClientSampleId :  
 ICVT102925

Tgt Ion: 43 Resp: 125522  
 Ion Ratio Lower Upper  
 43 100  
 86 10.0 7.8 11.6




Manual Integrations  
 APPROVED

Reviewed By :Amit Patel 10/30/2025  
 Supervised By :Mahesh Dadoda 10/30/2025




#24  
 1,1-Dichloroethane  
 Concen: 50.073 ug/l  
 RT: 5.662 min Scan# 709  
 Delta R.T. -0.000 min  
 Lab File: VT019199.D  
 Acq: 29 Oct 2025 12:42

Tgt Ion: 63 Resp: 152175  
 Ion Ratio Lower Upper  
 63 100  
 98 6.2 3.1 9.4  
 100 3.7 2.0 6.0








#25  
2-Butanone  
Concen: 257.837 ug/l  
RT: 6.667 min Scan# 81  
Delta R.T. 0.001 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVT102925




Tgt Ion: 43 Resp: 456774  
Ion Ratio Lower Upper  
43 100  
72 25.6 19.0 28.6

Manual Integrations  
APPROVED



Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025




#26  
2,2-Dichloropropane  
Concen: 45.233 ug/l  
RT: 6.655 min Scan# 878  
Delta R.T. -0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42



Tgt Ion: 77 Resp: 118617  
Ion Ratio Lower Upper  
77 100  
97 22.5 10.7 32.1






#27  
 cis-1,2-Dichloroethene  
 Concen: 52.651 ug/l  
 RT: 6.661 min Scan# 81  
 Delta R.T. 0.006 min  
 Lab File: VT019199.D  
 Acq: 29 Oct 2025 12:42

Instrument :

MSVOA\_T

ClientSampleId :

ICVVT102925






Tgt Ion: 96 Resp: 87305  
 Ion Ratio Lower Upper  
 96 100  
 61 150.4 0.0 309.0  
 98 63.7 0.0 128.4

Manual Integrations

APPROVED




Reviewed By :Amit Patel 10/30/2025

Supervised By :Mahesh Dadoda 10/30/2025



#28  
 Bromochloromethane  
 Concen: 53.987 ug/l  
 RT: 7.048 min Scan# 945  
 Delta R.T. 0.006 min  
 Lab File: VT019199.D  
 Acq: 29 Oct 2025 12:42

Tgt Ion: 49 Resp: 87685  
 Ion Ratio Lower Upper  
 49 100  
 129 0.2 0.0 0.0#  
 130 61.4 48.6 73.0





#29

Tetrahydrofuran

Concen: 283.927 ug/l

RT: 7.066 min Scan# 946

Delta R.T. 0.000 min

Lab File: VT019199.D

Acq: 29 Oct 2025 12:42

Instrument :

MSVOA\_T

ClientSampleId :

ICVT102925





Tgt Ion: 42 Resp: 292655  
 Ion Ratio Lower Upper  
 42 100  
 72 42.9 30.1 45.1  
 71 40.8 28.4 42.6

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/30/2025

Supervised By :Mahesh Dadoda 10/30/2025



#30

Chloroform

Concen: 51.508 ug/l




RT: 7.236 min Scan# 977


Delta R.T. 0.006 min

Lab File: VT019199.D

Acq: 29 Oct 2025 12:42


Tgt Ion: 83 Resp: 146555  
 Ion Ratio Lower Upper  
 83 100  
 85 61.9 50.6 76.0





#31  
Cyclohexane  
Concen: 42.453 ug/l  
RT: 7.536 min Scan# 1026  
Delta R.T. 0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVVT102925



Tgt Ion: 56 Resp: 9593  
Ion Ratio Lower Upper  
56 100  
69 30.4 26.1 39.1  
84 85.3 63.6 95.4




Manual Integrations  
APPROVED


Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025



#32  
1,1,1-Trichloroethane  
Concen: 49.622 ug/l  
RT: 7.442 min Scan# 1012  
Delta R.T. 0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42

Tgt Ion: 97 Resp: 112387  
Ion Ratio Lower Upper  
97 100  
99 62.4 50.5 75.7  
61 47.1 37.1 55.7









#33  
 1,2-Dichloroethane-d4  
 Concen: 52.505 ug/l  
 RT: 7.942 min Scan# 1095  
 Delta R.T. 0.007 min  
 Lab File: VT019199.D  
 Acq: 29 Oct 2025 12:42

Instrument :  
 MSVOA\_T  
 ClientSampleId :  
 ICVVT102925

Tgt Ion: 65 Resp: 121609  
 Ion Ratio Lower Upper  
 65 100  
 67 52.1 0.0 102.0




Manual Integrations  
 APPROVED

Reviewed By :Amit Patel 10/30/2025  
 Supervised By :Mahesh Dadoda 10/30/2025



#34  
 1,4-Difluorobenzene  
 Concen: 50.000 ug/l  
 RT: 8.547 min Scan# 1200  
 Delta R.T. -0.000 min  
 Lab File: VT019199.D  
 Acq: 29 Oct 2025 12:42


Tgt Ion:114 Resp: 246224  
 Ion Ratio Lower Upper  
 114 100  
 63 21.6 0.0 47.2  
 88 15.1 0.0 32.6





#35  
Dibromofluoromethane  
Concen: 52.628 ug/l  
RT: 7.471 min Scan# 1015  
Delta R.T. 0.006 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVT102925




Tgt Ion: 113 Resp: 81958  
Ion Ratio Lower Upper  
113 100  
111 104.6 81.6 122.4  
192 18.0 15.3 22.9

Manual Integrations  
APPROVED



Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025




#36  
1,1-Dichloropropene  
Concen: 46.887 ug/l  
RT: 7.683 min Scan# 1053  
Delta R.T. -0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42



Tgt Ion: 75 Resp: 82130  
Ion Ratio Lower Upper  
75 100  
110 31.4 16.1 48.2  
77 31.2 25.1 37.7





#37

Ethyl Acetate

Concen: 56.243 ug/l

RT: 6.772 min Scan# 89

Delta R.T. 0.006 min

Lab File: VT019199.D

Acq: 29 Oct 2025 12:42

Instrument :

MSVOA\_T

ClientSampleId :

ICVT102925

Tgt Ion: 43 Resp: 16771





| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 43  | 100   |       |       |
| 61  | 14.7  | 10.0  | 15.0  |
| 70  | 9.1   | 7.3   | 10.9  |

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/30/2025

Supervised By :Mahesh Dadoda 10/30/2025



#38

Carbon Tetrachloride

Concen: 48.952 ug/l

RT: 7.665 min Scan# 1050




Delta R.T. 0.006 min


Lab File: VT019199.D

Acq: 29 Oct 2025 12:42

Tgt Ion:117 Resp: 77966


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 117 | 100   |       |       |
| 119 | 94.3  | 75.0  | 112.6 |
| 121 | 30.4  | 23.3  | 34.9  |





#39  
Methylcyclohexane  
Concen: 45.754 ug/l  
RT: 9.093 min Scan# 1291  
Delta R.T. 0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVT102925



Tgt Ion: 83 Resp: 8855  
Ion Ratio Lower Upper  
83 100  
55 81.0 67.5 101.3  
98 51.6 39.9 59.9




Manual Integrations  
APPROVED

Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025




#40  
Benzene  
Concen: 49.562 ug/l  
RT: 7.953 min Scan# 1099  
Delta R.T. 0.006 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42

Tgt Ion: 78 Resp: 293881  
Ion Ratio Lower Upper  
78 100  
77 25.3 19.1 28.7











#41  
Methacrylonitrile  
Concen: 50.888 ug/l  
RT: 7.019 min Scan# 940  
Delta R.T. 0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42

Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVVT102925

Tgt Ion: 41 Resp: 7634  
Ion Ratio Lower Upper  
41 100  
39 70.5 49.8 74.6  
67 88.8 60.0 90.0  
52 37.9 26.6 40.0




Manual Integrations  
APPROVED


Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025



#42  
1,2-Dichloroethane  
Concen: 53.717 ug/l  
RT: 8.041 min Scan# 1114  
Delta R.T. 0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42


Tgt Ion: 62 Resp: 111929  
Ion Ratio Lower Upper  
62 100  
98 8.8 0.0 16.6





#43  
Isopropyl Acetate  
Concen: 54.759 ug/l  
RT: 8.094 min Scan# 1123  
Delta R.T. 0.006 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVT102925




Tgt Ion: 43 Resp: 25964  
Ion Ratio Lower Upper  
43 100  
61 24.3 14.8 22.2  
87 12.1 9.1 13.7

Manual Integrations  
APPROVED



Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025




#44  
Trichloroethene  
Concen: 48.482 ug/l  
RT: 8.823 min Scan# 1247  
Delta R.T. 0.006 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Tgt Ion:130 Resp: 56194  
Ion Ratio Lower Upper  
130 100  
95 99.7 0.0 207.8





#45  
1,2-Dichloropropane  
Concen: 51.381 ug/l  
RT: 9.140 min Scan# 1301  
Delta R.T. 0.006 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVVT102925



Tgt Ion: 63 Resp: 77519  
Ion Ratio Lower Upper  
63 100  
65 28.9 24.6 36.8




Manual Integrations  
APPROVED


Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025



#46  
Dibromomethane  
Concen: 53.227 ug/l  
RT: 9.240 min Scan# 1318  
Delta R.T. 0.006 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42


Tgt Ion: 93 Resp: 48331  
Ion Ratio Lower Upper  
93 100  
95 83.0 66.4 99.6  
174 87.1 77.3 115.9





#47  
Bromodichloromethane  
Concen: 52.999 ug/l  
RT: 9.457 min Scan# 1355  
Delta R.T. 0.006 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVT102925



Tgt Ion: 83 Resp: 10849  
Ion Ratio Lower Upper  
83 100  
85 61.5 50.4 75.6  
127 7.0 6.0 9.0




Manual Integrations  
APPROVED


Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025



#48  
Methyl methacrylate  
Concen: 54.093 ug/l  
RT: 9.228 min Scan# 1316  
Delta R.T. 0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42


Tgt Ion: 41 Resp: 115207  
Ion Ratio Lower Upper  
41 100  
69 83.3 56.9 85.3  
39 65.3 48.0 72.0





#49  
1,4-Dioxane  
Concen: 1116.760 ug/l  
RT: 9.234 min Scan# 1116  
Delta R.T. 0.006 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVVT102925




Tgt Ion: 88 Resp: 44673  
Ion Ratio Lower Upper  
88 100  
43 33.7 30.0 45.0  
58 68.2 57.2 85.8

Manual Integrations  
APPROVED



Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025




#50  
Toluene-d8  
Concen: 48.477 ug/l  
RT: 10.233 min Scan# 1487  
Delta R.T. 0.006 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Tgt Ion: 98 Resp: 306573  
Ion Ratio Lower Upper  
98 100  
100 66.2 53.5 80.3





#51  
4-Methyl-2-Pentanone  
Concen: 276.688 ug/l  
RT: 10.110 min Scan# 1466  
Delta R.T. 0.001 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVVT102925



Tgt Ion: 43 Resp: 790140  
Ion Ratio Lower Upper  
43 100  
58 40.6 30.5 45.7




Manual Integrations  
APPROVED


Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025



#52  
Toluene  
Concen: 48.815 ug/l  
RT: 10.304 min Scan# 1499  
Delta R.T. 0.006 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42


Tgt Ion: 92 Resp: 165395  
Ion Ratio Lower Upper  
92 100  
91 172.6 134.8 202.2





#53  
t-1,3-Dichloropropene  
Concen: 52.442 ug/l  
RT: 10.556 min Scan# 11  
Delta R.T. 0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVVT102925




Tgt Ion: 75 Resp: 11619  
Ion Ratio Lower Upper  
75 100  
77 29.8 24.2 36.4

Manual Integrations  
APPROVED



Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025




#54  
cis-1,3-Dichloropropene  
Concen: 52.367 ug/l  
RT: 9.945 min Scan# 1438  
Delta R.T. 0.006 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Tgt Ion: 75 Resp: 120149  
Ion Ratio Lower Upper  
75 100  
77 33.8 25.5 38.3  
39 56.3 48.8 73.2





#55  
1,1,2-Trichloroethane  
Concen: 53.334 ug/l  
RT: 10.762 min Scan# 1577  
Delta R.T. -0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVT102925



Tgt Ion: 97 Resp: 71272  
Ion Ratio Lower Upper  
97 100  
83 85.6 69.2 103.8  
85 55.9 43.8 65.6  
99 62.1 48.8 73.2




Manual Integrations  
APPROVED

Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025




#56  
Ethyl methacrylate  
Concen: 54.746 ug/l  
RT: 10.615 min Scan# 1552  
Delta R.T. 0.006 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42

Tgt Ion: 69 Resp: 140593  
Ion Ratio Lower Upper  
69 100  
41 68.5 62.6 93.8  
39 45.2 37.2 55.8








#57  
1,3-Dichloropropane  
Concen: 52.883 ug/l  
RT: 10.926 min Scan# 1605  
Delta R.T. 0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVT102925



Tgt Ion: 76 Resp: 125423  
Ion Ratio Lower Upper  
76 100  
78 32.1 25.9 38.9




Manual Integrations  
APPROVED


Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025



#58  
2-Chloroethyl Vinyl ether  
Concen: 267.903 ug/l  
RT: 9.786 min Scan# 1411  
Delta R.T. 0.006 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42


Tgt Ion: 63 Resp: 185157  
Ion Ratio Lower Upper  
63 100  
106 26.5 22.4 33.6





#59  
2-Hexanone  
Concen: 260.996 ug/l  
RT: 10.979 min Scan# 1614  
Delta R.T. 0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVT102925



Tgt Ion: 43 Resp: 564629  
Ion Ratio Lower Upper  
43 100  
58 57.0 26.1 78.3




Manual Integrations  
APPROVED


Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025



#60  
Dibromochloromethane  
Concen: 53.986 ug/l  
RT: 11.155 min Scan# 1644  
Delta R.T. 0.006 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42


Tgt Ion:129 Resp: 72737  
Ion Ratio Lower Upper  
129 100  
127 78.1 38.9 116.7





#61  
1,2-Dibromoethane  
Concen: 52.349 ug/l  
RT: 11.273 min Scan# 1664  
Delta R.T. 0.006 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVT102925



Tgt Ion:107 Resp: 65140  
Ion Ratio Lower Upper  
107 100  
109 96.1 75.6 113.4




Manual Integrations  
APPROVED


Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025



#62  
4-Bromofluorobenzene  
Concen: 48.398 ug/l  
RT: 12.859 min Scan# 1934  
Delta R.T. 0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42


Tgt Ion: 95 Resp: 96220  
Ion Ratio Lower Upper  
95 100  
174 79.4 0.0 160.6  
176 77.1 0.0 156.6





#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.725 min Scan# 117  
Delta R.T. 0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVVT102925



Tgt Ion:117 Resp: 19962  
Ion Ratio Lower Upper  
117 100  
82 59.4 45.1 67.7  
119 32.3 26.6 40.0




Manual Integrations  
APPROVED


Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025



#64  
Tetrachloroethene  
Concen: 47.070 ug/l  
RT: 10.844 min Scan# 1591  
Delta R.T. 0.006 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42


Tgt Ion:164 Resp: 42710  
Ion Ratio Lower Upper  
164 100  
166 126.4 100.9 151.3  
129 96.4 72.4 108.6  
131 85.5 67.9 101.9





#65  
Chlorobenzene  
Concen: 49.287 ug/l  
RT: 11.755 min Scan# 1761  
Delta R.T. -0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVVT102925



Tgt Ion:112 Resp: 17659  
Ion Ratio Lower Upper  
112 100  
114 31.1 25.8 38.8




Manual Integrations  
APPROVED


Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025



#66  
1,1,1,2-Tetrachloroethane  
Concen: 50.606 ug/l  
RT: 11.843 min Scan# 1761  
Delta R.T. 0.006 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42

Tgt Ion:131 Resp: 62079  
Ion Ratio Lower Upper  
131 100  
133 93.5 47.4 142.3  
119 66.2 31.7 95.1









#67  
Ethyl Benzene  
Concen: 46.761 ug/l  
RT: 11.843 min Scan# 1761  
Delta R.T. -0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42

Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVVT102925

Tgt Ion: 91 Resp: 313249  
Ion Ratio Lower Upper  
91 100  
106 30.8 24.8 37.2




Manual Integrations  
APPROVED


Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025



#68  
m/p-Xylenes  
Concen: 95.489 ug/l  
RT: 11.972 min Scan# 1783  
Delta R.T. 0.006 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42

Tgt Ion:106 Resp: 236279  
Ion Ratio Lower Upper  
106 100  
91 201.0 161.0 241.6









#69  
o-Xylene  
Concen: 48.287 ug/l  
RT: 12.342 min Scan# 1846  
Delta R.T. 0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42

Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVT102925

Tgt Ion:106 Resp: 118760  
Ion Ratio Lower Upper  
106 100  
91 215.5 106.5 319.5




Manual Integrations  
APPROVED


Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025



#70  
Styrene  
Concen: 49.670 ug/l  
RT: 12.360 min Scan# 1849  
Delta R.T. -0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42

Tgt Ion:104 Resp: 219989  
Ion Ratio Lower Upper  
104 100  
78 48.8 40.8 61.2  
103 55.5 43.8 65.6





#71

Bromoform

Concen: 55.844 ug/l

RT: 12.548 min Scan# 1881

Delta R.T. -0.000 min

Lab File: VT019199.D

Acq: 29 Oct 2025 12:42

Instrument :

MSVOA\_T

ClientSampleId :

ICVVT102925





Tgt Ion:173 Resp: 5238  
 Ion Ratio Lower Upper  
 173 100  
 175 49.1 24.0 72.0  
 254 0.0 0.0 0.0

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/30/2025

Supervised By :Mahesh Dadoda 10/30/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.929 min Scan# 2116

Delta R.T. -0.000 min


Lab File: VT019199.D

Acq: 29 Oct 2025 12:42

Tgt Ion:152 Resp: 98670  
 Ion Ratio Lower Upper  
 152 100  
 115 58.4 29.6 88.9  
 150 172.5 0.0 342.6











#73  
Isopropylbenzene  
Concen: 46.505 ug/l  
RT: 12.683 min Scan# 1904  
Delta R.T. 0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42

Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVVT102925

Tgt Ion:105 Resp: 315892  
Ion Ratio Lower Upper  
105 100  
120 25.5 13.5 40.4




Manual Integrations  
APPROVED


Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025



#74  
N-ethyl acetate  
Concen: 52.862 ug/l  
RT: 12.472 min Scan# 1868  
Delta R.T. 0.001 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42


Tgt Ion: 43 Resp: 181126  
Ion Ratio Lower Upper  
43 100  
70 43.9 31.3 46.9  
55 22.7 17.4 26.0  
61 24.4 18.1 27.1





#75  
1,1,2,2-Tetrachloroethane  
Concen: 52.108 ug/l  
RT: 12.977 min Scan# 1954  
Delta R.T. 0.006 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Instrument : MSVOA\_T  
ClientSampleId : ICVVT102925



Tgt Ion: 83 Resp: 121280  
Ion Ratio Lower Upper  
83 100  
131 8.8 5.2 15.6  
85 62.1 32.8 98.4




Manual Integrations  
APPROVED


Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025



#76  
1,2,3-Trichloropropane  
Concen: 51.189 ug/l m  
RT: 13.030 min Scan# 1963  
Delta R.T. -0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42

Tgt Ion: 75 Resp: 103936  
Ion Ratio Lower Upper  
75 100  
77 0.0 0.0 0.0









#77  
Bromobenzene  
Concen: 48.946 ug/l  
RT: 13.000 min Scan# 1957  
Delta R.T. 0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42

Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVT102925

Tgt Ion:156 Resp: 7054  
Ion Ratio Lower Upper  
156 100  
77 243.8 113.1 339.3  
158 97.3 48.0 144.0




Manual Integrations  
APPROVED


Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025



#78  
n-propylbenzene  
Concen: 46.314 ug/l  
RT: 13.077 min Scan# 1971  
Delta R.T. 0.006 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42

Tgt Ion: 91 Resp: 373758  
Ion Ratio Lower Upper  
91 100  
120 21.6 11.6 34.8





#79

2-Chlorotoluene

Concen: 47.160 ug/l

RT: 13.171 min Scan# 1986

Delta R.T. -0.000 min

Lab File: VT019199.D

Acq: 29 Oct 2025 12:42

Instrument :

MSVOA\_T

ClientSampleId :

ICVVT102925

Tgt Ion: 91 Resp: 218039

Ion Ratio Lower Upper

91 100





126 31.9 17.1 51.2

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/30/2025

Supervised By :Mahesh Dadoda 10/30/2025



#80

1,3,5-Trimethylbenzene

Concen: 46.838 ug/l

RT: 13.230 min Scan# 1997

Delta R.T. 0.001 min

Lab File: VT019199.D




Acq: 29 Oct 2025 12:42


Tgt Ion:105 Resp: 255548

Ion Ratio Lower Upper

105 100


120 50.0 25.5 76.5





#81  
trans-1,4-Dichloro-2-butene  
Concen: 55.825 ug/l  
RT: 12.742 min Scan# 1913  
Delta R.T. 0.006 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVT102925




Tgt Ion: 75 Resp: 48309  
Ion Ratio Lower Upper  
75 100  
53 96.6 86.0 129.0  
89 42.6 36.3 54.5

Manual Integrations  
APPROVED



Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025




#82  
4-Chlorotoluene  
Concen: 48.249 ug/l  
RT: 13.282 min Scan# 2006  
Delta R.T. 0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Tgt Ion: 91 Resp: 227000  
Ion Ratio Lower Upper  
91 100  
126 32.5 16.9 50.7





#83  
tert-Butylbenzene  
Concen: 47.610 ug/l  
RT: 13.529 min Scan# 2047  
Delta R.T. 0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVT102925




Tgt Ion:119 Resp: 240420  
Ion Ratio Lower Upper  
119 100  
91 63.2 32.9 98.6  
134 26.3 14.8 44.3

Manual Integrations  
APPROVED



Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025




#84  
1,2,4-Trimethylbenzene  
Concen: 46.724 ug/l  
RT: 13.582 min Scan# 2057  
Delta R.T. 0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42



Tgt Ion:105 Resp: 258900  
Ion Ratio Lower Upper  
105 100  
120 45.6 23.7 71.1









#85  
 sec-Butylbenzene  
 Concen: 46.734 ug/l  
 RT: 13.735 min Scan# 2082  
 Delta R.T. 0.006 min  
 Lab File: VT019199.D  
 Acq: 29 Oct 2025 12:42

Instrument :  
 MSVOA\_T  
 ClientSampleId :  
 ICVT102925

Tgt Ion:105 Resp: 34616  
 Ion Ratio Lower Upper  
 105 100  
 134 19.0 10.7 32.1




Manual Integrations  
 APPROVED


Reviewed By :Amit Patel 10/30/2025  
 Supervised By :Mahesh Dadoda 10/30/2025



#86  
 p-Isopropyltoluene  
 Concen: 45.267 ug/l  
 RT: 13.864 min Scan# 2105  
 Delta R.T. 0.000 min  
 Lab File: VT019199.D  
 Acq: 29 Oct 2025 12:42

Tgt Ion:119 Resp: 279948  
 Ion Ratio Lower Upper  
 119 100  
 134 25.8 13.4 40.1  
 91 22.2 11.7 35.1





#87

1,3-Dichlorobenzene

Concen: 47.618 ug/l

RT: 13.864 min Scan# 2104

Delta R.T. 0.000 min

Lab File: VT019199.D

Acq: 29 Oct 2025 12:42

Instrument :

MSVOA\_T

ClientSampleId :

ICVT102925

Tgt Ion:146 Resp: 14080

Ion Ratio Lower Upper

146 100

111 40.5 20.3 60.9





148 63.2 32.1 96.3

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/30/2025

Supervised By :Mahesh Dadoda 10/30/2025



#88

1,4-Dichlorobenzene

Concen: 47.789 ug/l

RT: 13.952 min Scan# 2120

Delta R.T. 0.000 min

Lab File: VT019199.D

Acq: 29 Oct 2025 12:42




Tgt Ion:146 Resp: 142349

Ion Ratio Lower Upper


146 100

111 38.8 19.3 57.8

148 64.4 31.8 95.4











#89  
n-Butylbenzene  
Concen: 47.433 ug/l  
RT: 14.234 min Scan# 2168  
Delta R.T. 0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42

Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVVT102925

Tgt Ion: 91 Resp: 270954  
Ion Ratio Lower Upper  
91 100  
92 52.8 27.1 81.2  
134 24.1 13.7 41.0




Manual Integrations  
APPROVED


Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025



#90  
Hexachloroethane  
Concen: 48.236 ug/l  
RT: 14.546 min Scan# 2221  
Delta R.T. 0.006 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42


Tgt Ion:117 Resp: 46022  
Ion Ratio Lower Upper  
117 100  
201 73.4 39.7 119.1





#91  
1,2-Dichlorobenzene  
Concen: 49.459 ug/l  
RT: 14.287 min Scan# 2176  
Delta R.T. 0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVT102925




Tgt Ion:146 Resp: 143274  
Ion Ratio Lower Upper  
146 100  
111 40.0 20.8 62.4  
148 63.6 32.6 97.8

Manual Integrations  
APPROVED



Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025




#92  
1,2-Dibromo-3-Chloropropane  
Concen: 54.900 ug/l  
RT: 14.986 min Scan# 2296  
Delta R.T. 0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Tgt Ion: 75 Resp: 27557  
Ion Ratio Lower Upper  
75 100  
155 82.1 42.9 128.7  
157 102.2 54.9 164.8





#93  
1,2,4-Trichlorobenzene  
Concen: 49.324 ug/l  
RT: 15.774 min Scan# 2430  
Delta R.T. 0.001 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVVT102925




Tgt Ion:180 Resp: 91890  
Ion Ratio Lower Upper  
180 100  
182 95.0 47.0 141.0  
145 30.2 15.0 45.0

Manual Integrations  
APPROVED



Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025




#94  
Hexachlorobutadiene  
Concen: 47.564 ug/l  
RT: 15.897 min Scan# 2451  
Delta R.T. 0.000 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Tgt Ion:225 Resp: 36550  
Ion Ratio Lower Upper  
225 100  
223 63.5 31.1 93.5  
227 62.0 31.9 95.9





#95  
Naphthalene  
Concen: 51.817 ug/l  
RT: 16.062 min Scan# 2479  
Delta R.T. 0.001 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42




Instrument :  
MSVOA\_T  
ClientSampleId :  
ICVT102925




Tgt Ion:128 Resp: 33803  
Ion Ratio Lower Upper  
128 100  
127 14.0 10.5 15.7  
129 11.0 8.8 13.2

Manual Integrations  
APPROVED

Reviewed By :Amit Patel 10/30/2025  
Supervised By :Mahesh Dadoda 10/30/2025



#96  
1,2,3-Trichlorobenzene  
Concen: 49.709 ug/l  
RT: 16.297 min Scan# 2519  
Delta R.T. 0.001 min  
Lab File: VT019199.D  
Acq: 29 Oct 2025 12:42



Tgt Ion:180 Resp: 91373  
Ion Ratio Lower Upper  
180 100  
182 95.3 47.6 142.8  
145 32.8 15.8 47.4

