

Data Path : Z:\VOASRV\HPCHEM1\MSVOA T\DATA\VT112818\
 Data File : VT019104.D
 Acq On : 28 Nov 2018 12:20
 Operator : SY/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA T/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_T
 ClientSampleId :
 VSTDICC100

Quant Time: Nov 29 06:16:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_T\METHODS\82T112818TCLP.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 29 06:13:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.45	168	578969	50.00	ug/l	0.00
7) 1,4-Difluorobenzene	8.38	114	1022138	50.00	ug/l	0.00
15) Chlorobenzene-d5	11.24	117	1023740	50.00	ug/l	0.00
18) 1,4-Dichlorobenzene-d4	13.16	152	546278	50.00	ug/l	0.00
System Monitoring Compounds						
6) 1,2-Dichloroethane-d4	7.82	65	1083641	95.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	190.54%	
8) Dibromofluoromethane	7.37	113	692951	105.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	210.64%	
13) Toluene-d8	9.91	98	3107159	107.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	215.70%	
14) 4-Bromofluorobenzene	12.23	95	1214319	112.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	224.20%	
Target Compounds						
					Qvalue	
2) Vinyl Chloride	2.11	62	1555785	88.141	ug/l	98
3) 1,1-Dichloroethene	3.57	96	1070353	88.144	ug/l	93
4) 2-Butanone	6.59	43	2616161	481.288	ug/l	97
5) Chloroform	7.13	83	1457391	91.656	ug/l	100
9) Carbon Tetrachloride	7.56	117	1173053	107.037	ug/l	98
10) Benzene	7.83	78	3721345	104.360	ug/l	98
11) 1,2-Dichloroethane	7.91	62	1285147	98.691	ug/l	100
12) Trichloroethene	8.64	130	741293	106.223	ug/l	100
16) Tetrachloroethene	10.45	164	618930	108.512	ug/l	99
17) Chlorobenzene	11.26	112	2391980	103.012	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA T\DATA\VT112818\
Data File : VT019104.D
Acq On : 28 Nov 2018 12:20
Operator : SY/MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA T/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_T
ClientSampleId :
VSTDICC100

Quant Time: Nov 29 06:16:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_T\METHODS\82T112818TCLP.M
Quant Title : SW846 8260
QLast Update : Thu Nov 29 06:13:05 2018
Response via : Initial Calibration

