

Data Path : Z:\VOASRV\HPCHEM1\MSVOA T\DATA\VT112818\
 Data File : VT019114.D
 Acq On : 28 Nov 2018 19:50
 Operator : SY/MD
 Sample : MDL6
 Misc : 5.0mL/MSVOA T/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_T
 ClientSampleId :
 MDL6

Quant Time: Nov 29 06:31:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_T\METHODS\82T112818TCLP.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 29 06:21:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.45	168	257041	50.00	ug/l	0.00
7) 1,4-Difluorobenzene	8.38	114	540326	50.00	ug/l	0.00
15) Chlorobenzene-d5	11.24	117	501100	50.00	ug/l	0.00
18) 1,4-Dichlorobenzene-d4	13.16	152	237031	50.00	ug/l	0.00
System Monitoring Compounds						
6) 1,2-Dichloroethane-d4	7.82	65	337572	66.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	133.70%	
8) Dibromofluoromethane	7.37	113	184244	52.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.96%	
13) Toluene-d8	9.91	98	831911	54.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.26%	
14) 4-Bromofluorobenzene	12.23	95	298283	52.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.18%	
Target Compounds						
					Qvalue	
2) Vinyl Chloride	2.10	62	25843	3.298	ug/l	97
3) 1,1-Dichloroethene	3.57	96	20993	3.894	ug/l	100
4) 2-Butanone	6.59	43	37488	15.534	ug/l	95
5) Chloroform	7.13	83	21657	3.068	ug/l	91
9) Carbon Tetrachloride	7.56	117	15098	2.606	ug/l	91
10) Benzene	7.82	78	47124	2.500	ug/l	99
11) 1,2-Dichloroethane	7.91	62	20519	2.981	ug/l	97
12) Trichloroethene	8.64	130	8472	2.297	ug/l	91
16) Tetrachloroethene	10.45	164	5713	2.046	ug/l #	88
17) Chlorobenzene	11.26	112	27089	2.383	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA T\DATA\VT112818\
Data File : VT019114.D
Acq On : 28 Nov 2018 19:50
Operator : SY/MD
Sample : MDL6
Misc : 5.0mL/MSVOA T/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_T
ClientSampleId :
MDL6

Quant Time: Nov 29 06:31:07 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_T\METHODS\82T112818TCLP.M
Quant Title : SW846 8260
QLast Update : Thu Nov 29 06:21:27 2018
Response via : Initial Calibration

