

Data Path : Z:\VOASRV\HPCHEM1\MSVOA T\DATA\VT112818\
 Data File : VT019117.D
 Acq On : 28 Nov 2018 21:19
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA T/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_T
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Nov 29 06:31:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_T\METHODS\82T112818TCLP.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 29 06:21:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.45	168	316754	50.00	ug/l	0.00
7) 1,4-Difluorobenzene	8.38	114	599830	50.00	ug/l	0.00
15) Chlorobenzene-d5	11.24	117	607967	50.00	ug/l	0.00
18) 1,4-Dichlorobenzene-d4	13.16	152	315479	50.00	ug/l	0.00

System Monitoring Compounds

6) 1,2-Dichloroethane-d4	7.82	65	366400	58.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	117.76%	
8) Dibromofluoromethane	7.37	113	214686	55.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	111.20%	
13) Toluene-d8	9.91	98	945502	55.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	111.86%	
14) 4-Bromofluorobenzene	12.23	95	366722	57.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	115.38%	

Target Compounds

					Qvalue
2) Vinyl Chloride	2.11	62	537008	55.609	ug/l 99
3) 1,1-Dichloroethene	3.56	96	358689	53.996	ug/l # 82
4) 2-Butanone	6.57	43	850201	285.888	ug/l 97
5) Chloroform	7.13	83	478690	55.027	ug/l 100
9) Carbon Tetrachloride	7.54	117	357146	55.532	ug/l 99
10) Benzene	7.82	78	1121654	53.601	ug/l 100
11) 1,2-Dichloroethane	7.91	62	443463	58.032	ug/l 96
12) Trichloroethene	8.64	130	202236	49.382	ug/l 91
16) Tetrachloroethene	10.45	164	168638	49.785	ug/l 91
17) Chlorobenzene	11.26	112	686099	49.754	ug/l 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA T\DATA\VT112818\
Data File : VT019117.D
Acq On : 28 Nov 2018 21:19
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA T/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_T
ClientSampleId :
VSTDCCC050EC

Quant Time: Nov 29 06:31:38 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_T\METHODS\82T112818TCLP.M
Quant Title : SW846 8260
QLast Update : Thu Nov 29 06:21:27 2018
Response via : Initial Calibration

