

Data Path : Z:\voasrv\HPCHEM1\MSVOA T\Data\VT112818\
 Data File : VT019117.D
 Acq On : 28 Nov 2018 21:19
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA T/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_T
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 29 06:31:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_T\METHODS\82T112818TCLP.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 29 06:21:27 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

1 I	Pentafluorobenzene	1.000	1.000	0.0	63	0.00
2 C	Vinyl Chloride	1.524	1.695	-11.2#	70	0.00
3 CM	1,1-Dichloroethene	1.049	1.132	-7.9#	70	0.00
4 T	2-Butanone	0.469	0.537	-14.5	68	0.00
5 C	Chloroform	1.373	1.511	-10.1#	70	0.00
6 S	1,2-Dichloroethane-d4	0.982	1.157	-17.8	73	0.00
7 I	1,4-Difluorobenzene	1.000	1.000	0.0	64	0.00
8 S	Dibromofluoromethane	0.322	0.358	-11.2	70	0.00
9 T	Carbon Tetrachloride	0.536	0.595	-11.0	68	0.00
10 TM	Benzene	1.744	1.870	-7.2	66	-0.01
11 TM	1,2-Dichloroethane	0.637	0.739	-16.0	73	0.00
12 TM	Trichloroethene	0.341	0.337	1.2	61	0.00
13 S	Toluene-d8	1.409	1.576	-11.9	69	0.00
14 S	4-Bromofluorobenzene	0.530	0.611	-15.3	69	0.00
15 I	Chlorobenzene-d5	1.000	1.000	0.0	65	0.00
16 T	Tetrachloroethene	0.279	0.277	0.7	63	0.00
17 PM	Chlorobenzene	1.134	1.129	0.4	64	0.00
18 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	64	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 3