

Data Path : Z:\voasrv\HPCHEM1\MSVOA T\Data\VT112818\
 Data File : VT019117.D
 Acq On : 28 Nov 2018 21:19
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA T/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_T
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 29 06:31:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_T\METHODS\82T112818TCLP.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 29 06:21:27 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

1 I	Pentafluorobenzene	50.000	50.000	0.0	63	0.00
2 C	Vinyl Chloride	50.000	55.609	-11.2#	70	0.00
3 CM	1,1-Dichloroethene	50.000	53.996	-8.0#	70	0.00
4 T	2-Butanone	250.000	285.888	-14.4	68	0.00
5 C	Chloroform	50.000	55.027	-10.1#	70	0.00
6 S	1,2-Dichloroethane-d4	50.000	58.882	-17.8	73	0.00
7 I	1,4-Difluorobenzene	50.000	50.000	0.0	64	0.00
8 S	Dibromofluoromethane	50.000	55.605	-11.2	70	0.00
9 T	Carbon Tetrachloride	50.000	55.532	-11.1	68	0.00
10 TM	Benzene	50.000	53.601	-7.2	66	-0.01
11 TM	1,2-Dichloroethane	50.000	58.032	-16.1	73	0.00
12 TM	Trichloroethene	50.000	49.382	1.2	61	0.00
13 S	Toluene-d8	50.000	55.927	-11.9	69	0.00
14 S	4-Bromofluorobenzene	50.000	57.689	-15.4	69	0.00
15 I	Chlorobenzene-d5	50.000	50.000	0.0	65	0.00
16 T	Tetrachloroethene	50.000	49.785	0.4	63	0.00
17 PM	Chlorobenzene	50.000	49.754	0.5	64	0.00
18 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	64	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 3