

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU011119\
 Data File : VU029132.D
 Acq On : 11 Jan 2019 10:14
 Operator : MD/SY
 Sample : VU0111WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK49

Quant Time: Jan 11 22:18:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR122818WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jan 11 22:10:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	136130	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	129181	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	55001	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	42333	5.81	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	116.20%
7) Chloroethane-d5	1.68	69	36566	5.75	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	115.00%
11) 1,1-Dichloroethene-d2	2.27	63	58937	3.82	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.40%
20) 2-Butanone-d5	4.22	46	119858	62.95	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	125.90%
24) Chloroform-d	4.64	84	77168	5.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.00%
26) 1,2-Dichloroethane-d4	5.31	65	39836	5.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.20%
32) Benzene-d6	5.34	84	166767	5.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.40%
36) 1,2-Dichloropropane-d6	6.33	67	55695	5.77	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	115.40%
41) Toluene-d8	7.56	98	143782	5.14	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.80%
43) trans-1,3-Dichloropropene-	7.85	79	20707	5.70	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	114.00%
46) 2-Hexanone-d5	8.31	63	102037	61.66	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	123.32%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	42309	5.51	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	110.20%
64) 1,2-Dichlorobenzene-d4	11.86	152	51917	5.64	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.80%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU011119\
Data File : VU029132.D
Acq On : 11 Jan 2019 10:14
Operator : MD/SY
Sample : VU0111WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK49

Quant Time: Jan 11 22:18:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR122818WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jan 11 22:10:09 2019
Response via : Initial Calibration

