

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU011119\
 Data File : VU029140.D
 Acq On : 11 Jan 2019 14:18
 Operator : JC/SP
 Sample : K1111-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Jan 11 22:54:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR122818WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jan 11 22:10:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	114359	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	114519	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	48232	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	37180	6.08	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	121.60%
7) Chloroethane-d5	1.68	69	32110	6.01	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	120.20%
11) 1,1-Dichloroethene-d2	2.27	63	51953	4.01	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.20%
20) 2-Butanone-d5	4.21	46	102149	63.86	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	127.72%
24) Chloroform-d	4.65	84	72272	6.07	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	121.40%
26) 1,2-Dichloroethane-d4	5.31	65	36698	6.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	121.00%
32) Benzene-d6	5.34	84	146897	5.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.80%
36) 1,2-Dichloropropane-d6	6.33	67	51274	5.99	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	119.80%
41) Toluene-d8	7.56	98	127889	5.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.20%
43) trans-1,3-Dichloropropene-	7.85	79	17933	5.57	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	111.40%
46) 2-Hexanone-d5	8.31	63	84610	57.67	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	115.34%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	38495	5.65	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	113.00%
64) 1,2-Dichlorobenzene-d4	11.86	152	45871	5.69	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.80%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU011119\
Data File : VU029140.D
Acq On : 11 Jan 2019 14:18
Operator : JC/SP
Sample : K1111-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Jan 11 22:54:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR122818WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jan 11 22:10:09 2019
Response via : Initial Calibration

