

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU010219\
 Data File : VU028986.D
 Acq On : 02 Jan 2019 13:22
 Operator : MD/SY
 Sample : VIBLK70
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK70

Quant Time: Jan 06 21:48:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR122818WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jan 03 03:47:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	95969	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	89611	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	42010	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	31137	6.06	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	121.20%
7) Chloroethane-d5	1.68	69	25516	5.69	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	113.80%
11) 1,1-Dichloroethene-d2	2.27	63	42825	3.94	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.80%
20) 2-Butanone-d5	4.19	46	78430	58.43	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	116.86%
24) Chloroform-d	4.65	84	55326	5.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.80%
26) 1,2-Dichloroethane-d4	5.31	65	26960	5.29	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.80%
32) Benzene-d6	5.34	84	117122	5.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	111.80%
36) 1,2-Dichloropropane-d6	6.33	67	38298	5.72	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	114.40%
41) Toluene-d8	7.56	98	104688	5.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.00%
43) trans-1,3-Dichloropropene-	7.85	79	14357	5.70	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	114.00%
46) 2-Hexanone-d5	8.31	63	71052	61.89	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	123.78%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	29566	5.55	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	111.00%
64) 1,2-Dichlorobenzene-d4	11.86	152	38623	5.50	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU010219\
Data File : VU028986.D
Acq On : 02 Jan 2019 13:22
Operator : MD/SY
Sample : VIBLK70
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK70

Quant Time: Jan 06 21:48:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR122818WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jan 03 03:47:56 2019
Response via : Initial Calibration

