

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU010219\
 Data File : VU028987.D
 Acq On : 02 Jan 2019 13:46
 Operator : MD/SY
 Sample : J6593-01DL 25X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BE803DL

Quant Time: Jan 03 05:05:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR122818WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jan 03 03:47:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	96362	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	91561	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	42396	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	30558	5.93	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	118.60%
7) Chloroethane-d5	1.68	69	25279	5.62	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	112.40%
11) 1,1-Dichloroethene-d2	2.27	63	41998	3.85	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.00%
20) 2-Butanone-d5	4.19	46	76293	56.60	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.20%
24) Chloroform-d	4.65	84	53106	5.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.00%
26) 1,2-Dichloroethane-d4	5.31	65	27242	5.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.60%
32) Benzene-d6	5.34	84	116266	5.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.60%
36) 1,2-Dichloropropane-d6	6.33	67	39457	5.77	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	115.40%
41) Toluene-d8	7.56	98	103581	5.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.60%
43) trans-1,3-Dichloropropene-	7.85	79	14234	5.53	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	110.60%
46) 2-Hexanone-d5	8.31	63	70609	60.20	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	120.40%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	29074	5.34	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.80%
64) 1,2-Dichlorobenzene-d4	11.86	152	38402	5.41	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.20%

Target Compounds					Ovalue
8) Chloroethane	1.70	64	43767	7.971 ug/L	96
33) Benzene	5.39	78	346736	11.843 ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU010219\
Data File : VU028987.D
Acq On : 02 Jan 2019 13:46
Operator : MD/SY
Sample : J6593-01DL 25X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BE803DL

Quant Time: Jan 03 05:05:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR122818WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jan 03 03:47:56 2019
Response via : Initial Calibration

