

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU010319\
 Data File : VU029000.D
 Acq On : 03 Jan 2019 16:18
 Operator : MD/SY
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ICVVU010319

Quant Time: Jan 04 03:52:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 04 02:22:53 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	107	0.00
2 T	Dichlorodifluoromethane	0.496	0.481	3.0	104	0.00
3 P	Chloromethane	0.887	0.746	15.9	100	0.00
4 C	Vinyl Chloride	0.731	0.700	4.2#	103	0.00
5 T	Bromomethane	0.437	0.404	7.6	106	0.00
6 T	Chloroethane	0.435	0.418	3.9	105	0.00
7 T	Trichlorofluoromethane	0.796	0.756	5.0	103	0.00
8 T	Diethyl Ether	0.383	0.363	5.2	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.533	0.522	2.1	117	0.00
10 T	Methyl Iodide	0.501	0.513	-2.4	111	0.00
11 T	Tert butyl alcohol	0.157	0.145	7.6	102	0.00
12 CM	1,1-Dichloroethene	0.519	0.517	0.4#	116	0.00
13 T	Acrolein	0.118	0.105	11.0	98	0.00
14 T	Allyl chloride	0.951	0.956	-0.5	103	0.00
15 T	Acrylonitrile	0.385	0.372	3.4	101	0.00
16 T	Acetone	0.389	0.359	7.7	103	0.00
17 T	Carbon Disulfide	1.930	1.652	14.4	103	0.00
18 T	Methyl Acetate	0.905	0.854	5.6	103	0.00
19 T	Methyl tert-butyl Ether	1.823	1.756	3.7	103	0.00
20 T	Methylene Chloride	0.661	0.621	6.1	104	0.00
21 T	trans-1,2-Dichloroethene	0.602	0.569	5.5	103	0.00
22 T	Diisopropyl ether	1.880	1.841	2.1	103	0.00
23 T	Vinyl Acetate	1.537	1.543	-0.4	104	0.00
24 P	1,1-Dichloroethane	1.086	1.056	2.8	103	0.00
25 T	2-Butanone	0.526	0.517	1.7	102	0.00
26 T	2,2-Dichloropropane	0.877	0.867	1.1	105	0.00
27 T	cis-1,2-Dichloroethene	0.659	0.653	0.9	105	0.00
28 T	Bromochloromethane	0.465	0.453	2.6	101	0.00
29 T	Tetrahydrofuran	0.319	0.309	3.1	101	0.00
30 C	Chloroform	1.034	1.006	2.7#	103	0.00
31 T	Cyclohexane	1.201	1.020	15.1	106	0.00
32 T	1,1,1-Trichloroethane	0.854	0.812	4.9	101	0.00
33 S	1,2-Dichloroethane-d4	0.618	0.608	1.6	104	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	104	0.00
35 S	Dibromofluoromethane	0.324	0.324	0.0	105	0.00
36 T	1,1-Dichloropropene	0.500	0.479	4.2	104	0.00
37 T	Ethyl Acetate	0.549	0.537	2.2	102	0.00
38 T	Carbon Tetrachloride	0.417	0.411	1.4	102	0.00
39 T	Methylcyclohexane	0.614	0.607	1.1	106	0.00
40 TM	Benzene	1.559	1.513	3.0	103	0.00
41 T	Methacrylonitrile	0.294	0.301	-2.4	105	0.00
42 TM	1,2-Dichloroethane	0.468	0.467	0.2	102	0.00
43 T	Isopropyl Acetate	0.842	0.853	-1.3	104	0.00
44 TM	Trichloroethene	0.386	0.370	4.1	103	0.00
45 C	1,2-Dichloropropane	0.412	0.400	2.9#	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU010319\
 Data File : VU029000.D
 Acq On : 03 Jan 2019 16:18
 Operator : MD/SY
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ICVVU010319

Quant Time: Jan 04 03:52:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 04 02:22:53 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.263	0.254	3.4	102	0.00
47 T	Bromodichloromethane	0.485	0.480	1.0	105	0.00
48 T	Methyl methacrylate	0.410	0.420	-2.4	101	0.00
49 T	1,4-Dioxane	0.010	0.009	10.0	99	0.00
50 S	Toluene-d8	1.280	1.272	0.6	104	0.00
51 T	4-Methyl-2-Pentanone	0.573	0.562	1.9	101	0.00
52 CM	Toluene	0.949	0.924	2.6#	103	0.00
53 T	t-1,3-Dichloropropene	0.582	0.565	2.9	103	0.00
54 T	cis-1,3-Dichloropropene	0.629	0.617	1.9	103	0.00
55 T	1,1,2-Trichloroethane	0.380	0.369	2.9	102	0.00
56 T	Ethyl methacrylate	0.586	0.595	-1.5	104	0.00
57 T	1,3-Dichloropropane	0.651	0.643	1.2	104	0.00
58 T	2-Chloroethyl Vinyl ether	0.258	0.269	-4.3	106	0.00
59 T	2-Hexanone	0.450	0.448	0.4	103	0.00
60 T	Dibromochloromethane	0.377	0.365	3.2	102	0.00
61 T	1,2-Dibromoethane	0.412	0.398	3.4	102	0.00
62 S	4-Bromofluorobenzene	0.460	0.458	0.4	106	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	104	0.00
64 T	Tetrachloroethene	0.360	0.359	0.3	103	0.00
65 PM	Chlorobenzene	1.121	1.079	3.7	104	0.00
66 T	1,1,1,2-Tetrachloroethane	0.364	0.366	-0.5	105	0.00
67 C	Ethyl Benzene	1.871	1.878	-0.4#	103	0.00
68 T	m/p-Xylenes	0.727	0.729	-0.3	104	0.00
69 T	o-Xylene	0.701	0.707	-0.9	104	0.00
70 T	Styrene	1.148	1.172	-2.1	103	0.00
71 P	Bromoform	0.324	0.323	0.3	101	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
73 T	Isopropylbenzene	3.701	3.619	2.2	104	0.00
74 T	N-amyl acetate	1.595	1.602	-0.4	104	0.00
75 P	1,1,2,2-Tetrachloroethane	1.483	1.402	5.5	103	0.00
76 T	1,2,3-Trichloropropane	1.225	1.183	3.4	102	0.00
77 T	Bromobenzene	0.964	0.929	3.6	104	0.00
78 T	n-propylbenzene	4.361	4.325	0.8	104	0.00
79 T	2-Chlorotoluene	2.646	2.521	4.7	103	0.00
80 T	1,3,5-Trimethylbenzene	3.091	3.036	1.8	104	0.00
81 T	trans-1,4-Dichloro-2-butene	0.475	0.416	12.4	101	0.00
82 T	4-Chlorotoluene	3.034	2.910	4.1	103	0.00
83 T	tert-Butylbenzene	3.046	2.961	2.8	104	0.00
84 T	1,2,4-Trimethylbenzene	3.086	3.116	-1.0	105	0.00
85 T	sec-Butylbenzene	3.684	3.644	1.1	104	0.00
86 T	p-Isopropyltoluene	3.139	3.174	-1.1	105	0.00
87 T	1,3-Dichlorobenzene	1.769	1.687	4.6	104	0.00
88 T	1,4-Dichlorobenzene	1.826	1.682	7.9	104	0.00
89 T	n-Butylbenzene	2.802	2.907	-3.7	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU010319\
Data File : VU029000.D
Acq On : 03 Jan 2019 16:18
Operator : MD/SY
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ICVVU010319

Quant Time: Jan 04 03:52:48 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
Quant Title : SW846 8260
QLast Update : Fri Jan 04 02:22:53 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.528	0.489	7.4	101	0.00
91 T	1,2-Dichlorobenzene	1.782	1.675	6.0	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.285	0.280	1.8	99	0.00
93 T	1,2,4-Trichlorobenzene	1.137	1.132	0.4	105	0.00
94 T	Hexachlorobutadiene	0.520	0.507	2.5	108	0.00
95 T	Naphthalene	3.869	3.726	3.7	102	0.00
96 T	1,2,3-Trichlorobenzene	1.209	1.133	6.3	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6