

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU010421\
 Data File : VU041846.D
 Acq On : 04 Jan 2021 16:14
 Operator : SY/MD
 Sample : L5254-05
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Jan 05 03:19:23 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jan 05 01:00:34 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	172651	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	164008	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	72567	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	52509	3.89	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.80%
7) Chloroethane-d5	1.92	69	52388	5.01	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.20%
11) 1,1-Dichloroethene-d2	2.57	63	81813	3.29	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.80%
20) 2-Butanone-d5	4.65	46	175550	45.28	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	90.56%
24) Chloroform-d	5.08	84	110137	4.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.20%
26) 1,2-Dichloroethane-d4	5.71	65	61361	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.60%
32) Benzene-d6	5.74	84	214829	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.40%
36) 1,2-Dichloropropane-d6	6.70	67	69019	4.69	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.80%
41) Toluene-d8	7.90	98	168577	3.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.60%
43) trans-1,3-Dichloropropene-	8.18	79	24845	3.75	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.00%
46) 2-Hexanone-d5	8.64	63	133773	36.96	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	73.92%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	55991	4.19	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.80%
64) 1,2-Dichlorobenzene-d4	12.19	152	65819	5.10	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU010421\
Data File : VU041846.D
Acq On : 04 Jan 2021 16:14
Operator : SY/MD
Sample : L5254-05
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Jan 05 03:19:23 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jan 05 01:00:34 2021
Response via : Initial Calibration

