


Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU010421\
Data File : VU041839.D
Acq On : 04 Jan 2021 12:33
Operator : SY/MD
Sample : L5254-04
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
B0AH6

Manual Integrations
APPROVED

SAM
1/7/2021 5:22:20 PM

Quant Time: Jan 05 01:42:18 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jan 05 01:00:34 2021
Response via : Initial Calibration



Quantitation Report (Qedit)




Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU010421\
 Data File : VU041839.D
 Acq On : 04 Jan 2021 12:33
 Operator : SY/MD
 Sample : L5254-04
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 B0AH6

Manual Integrations
 APPROVED

SAM
 1/7/2021 5:22:20 PM

Quant Time: Jan 05 01:01:22 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jan 05 01:00:34 2021
 Response via : Initial Calibration



(13) Acetone (T)
 2.668min (+0.006) 0.66ug/L
 response 1826

Ion	Exp%	Act%
43.05	100	100
58.05	36.60	38.39
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU010421\
Data File : VU041839.D
Acq On : 04 Jan 2021 12:33
Operator : SY/MD
Sample : L5254-04
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
B0AH6

Manual Integrations
APPROVED

SAM
1/7/2021 5:22:20 PM

Quant Time: Jan 05 01:01:22 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jan 05 01:00:34 2021
Response via : Initial Calibration



TIC: VU041839.D

(13) Acetone (T)

2.668min (+0.006) 1.21ug/L m

response 3356

Ion	Exp%	Act%
43.05	100	100
58.05	36.60	20.89
0.00	0.00	0.00
0.00	0.00	0.00

2 mg
01/08/21

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU010421\
 Data File : VU041839.D
 Acq On : 04 Jan 2021 12:33
 Operator : SY/MD
 Sample : L5254-04
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 B0AH6

Manual Integrations
 APPROVED

SAM
 1/7/2021 5:22:20 PM

Quant Time: Jan 05 01:42:18 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jan 05 01:00:34 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev (Min)
1) 1,4-Difluorobenzene	6.26	114	194852	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	195072	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	80120	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	50468	3.31	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	66.20%
7) Chloroethane-d5	1.92	69	48584	4.12	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.40%
11) 1,1-Dichloroethene-d2	2.57	63	70801	2.52	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	50.40%#
20) 2-Butanone-d5	4.65	46	159443	36.44	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	72.88%
24) Chloroform-d	5.08	84	100848	3.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.20%
26) 1,2-Dichloroethane-d4	5.71	65	56136	3.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.80%
32) Benzene-d6	5.74	84	195506	3.50	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	70.00%
36) 1,2-Dichloropropane-d6	6.70	67	65710	3.75	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	75.00%
41) Toluene-d8	7.90	98	159398	3.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	63.20%#
43) trans-1,3-Dichloropropene-	8.18	79	25659	3.26	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	65.20%
46) 2-Hexanone-d5	8.64	63	128031	29.74	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	59.48%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	49830	3.13	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	62.60%#
64) 1,2-Dichlorobenzene-d4	12.19	152	55341	3.88	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	77.60%#

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Acetone	2.67	43	3356m	1.206	ug/L	
42) Toluene	7.97	91	13229	0.205	ug/L	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

2 mD
 94
 01/08/21