

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU010421\
 Data File : VU041840.D
 Acq On : 04 Jan 2021 12:55
 Operator : SY/MD
 Sample : L5254-02
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 B0AH4

Quant Time: Jan 05 01:51:18 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jan 05 01:00:34 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	145360	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	146835	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	65419	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	57281	5.04	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.80%
7) Chloroethane-d5	1.92	69	55072	6.26	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	125.20%
11) 1,1-Dichloroethene-d2	2.58	63	81240	3.88	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.60%
20) 2-Butanone-d5	4.66	46	162665	49.83	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.66%
24) Chloroform-d	5.08	84	105672	5.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.60%
26) 1,2-Dichloroethane-d4	5.72	65	59805	5.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.40%
32) Benzene-d6	5.74	84	204649	4.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.40%
36) 1,2-Dichloropropane-d6	6.70	67	66165	5.02	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.40%
41) Toluene-d8	7.90	98	169992	4.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.60%
43) trans-1,3-Dichloropropene-	8.19	79	25842	4.36	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.20%
46) 2-Hexanone-d5	8.64	63	136419	42.10	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.20%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	56749	4.74	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.80%
64) 1,2-Dichlorobenzene-d4	12.19	152	60547	5.21	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU010421\
Data File : VU041840.D
Acq On : 04 Jan 2021 12:55
Operator : SY/MD
Sample : L5254-02
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
B0AH4

Quant Time: Jan 05 01:51:18 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jan 05 01:00:34 2021
Response via : Initial Calibration

