

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU010519\
 Data File : VU029035.D
 Acq On : 04 Jan 2019 18:18
 Operator : MD/SY
 Sample : K1055-03
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-3

Quant Time: Jan 05 04:27:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 04 02:22:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	110835	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.88	114	188849	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	172867	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	67517	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.30	65	74543	54.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.82%	
35) Dibromofluoromethane	4.88	113	64326	52.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.10%	
50) Toluene-d8	7.56	98	238163	49.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.56%	
62) 4-Bromofluorobenzene	10.31	95	73860	42.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.06%	
Target Compounds						
19) Methyl tert-butyl Ether	3.00	73	1925	0.476	ug/l #	89
27) cis-1,2-Dichloroethene	4.22	96	2445	1.674	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU010519\
Data File : VU029035.D
Acq On : 04 Jan 2019 18:18
Operator : MD/SY
Sample : K1055-03
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-3

Quant Time: Jan 05 04:27:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
Quant Title : SW846 8260
QLast Update : Fri Jan 04 02:22:53 2019
Response via : Initial Calibration

