

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU010519\
 Data File : VU029040.D
 Acq On : 04 Jan 2019 20:19
 Operator : MD/SY
 Sample : K1053-04 10X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COMP-2-DRUM

Quant Time: Jan 05 04:43:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 04 02:22:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	114484	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.88	114	193840	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	179506	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	73882	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	75297	53.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.40%	
35) Dibromofluoromethane	4.88	113	65892	52.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.88%	
50) Toluene-d8	7.56	98	242249	48.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.66%	
62) 4-Bromofluorobenzene	10.30	95	77060	43.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.46%	
Target Compounds						
16) Acetone	2.32	43	3559	3.992	ug/l	90
18) Methyl Acetate	2.62	43	1779	0.858	ug/l #	83
25) 2-Butanone	4.26	43	34432	28.594	ug/l	94
40) Benzene	5.39	78	2584	0.428	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU010519\
Data File : VU029040.D
Acq On : 04 Jan 2019 20:19
Operator : MD/SY
Sample : K1053-04 10X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
COMP-2-DRUM

Quant Time: Jan 05 04:43:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
Quant Title : SW846 8260
QLast Update : Fri Jan 04 02:22:53 2019
Response via : Initial Calibration

