

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU010519\
 Data File : VU029041.D
 Acq On : 04 Jan 2019 20:43
 Operator : MD/SY
 Sample : K1053-07 10X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COMP-3-DRUM

Quant Time: Jan 05 04:46:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 04 02:22:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	111091	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.88	114	192419	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	173947	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	68889	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	74292	54.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.20%	
35) Dibromofluoromethane	4.88	113	65303	52.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.70%	
50) Toluene-d8	7.56	98	238758	48.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.98%	
62) 4-Bromofluorobenzene	10.30	95	74396	42.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.08%	
Target Compounds						
16) Acetone	2.32	43	2593	2.997	ug/l	94
25) 2-Butanone	4.27	43	23472	20.088	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU010519\
Data File : VU029041.D
Acq On : 04 Jan 2019 20:43
Operator : MD/SY
Sample : K1053-07 10X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
COMP-3-DRUM

Quant Time: Jan 05 04:46:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
Quant Title : SW846 8260
QLast Update : Fri Jan 04 02:22:53 2019
Response via : Initial Calibration

