

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU010519\
 Data File : VU029042.D
 Acq On : 04 Jan 2019 21:08
 Operator : MD/SY
 Sample : K1059-03 10X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BUR-COMP-1

Quant Time: Jan 05 04:49:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 04 02:22:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	110689	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.88	114	188663	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	170991	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	65408	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	74852	54.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.40%	
35) Dibromofluoromethane	4.88	113	62886	51.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.84%	
50) Toluene-d8	7.56	98	236344	48.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.90%	
62) 4-Bromofluorobenzene	10.30	95	71928	41.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.90%	
Target Compounds						
16) Acetone	2.32	43	1887	2.189	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU010519\
Data File : VU029042.D
Acq On : 04 Jan 2019 21:08
Operator : MD/SY
Sample : K1059-03 10X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BUR-COMP-1

Quant Time: Jan 05 04:49:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
Quant Title : SW846 8260
QLast Update : Fri Jan 04 02:22:53 2019
Response via : Initial Calibration

