


Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU010521\
 Data File : VU041848.D
 Acq On : 05 Jan 2021 09:16
 Operator : SY/MD
 Sample : BFB36
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB36

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SFAMUTR122820WMA.M
 Title : TRACE VOA SFAM1.0
 Last Update : Wed Jan 06 06:32:42 2021



AutoFind: Scans 2889, 2890, 2891; Background Corrected with Scan 2875

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.8	19672	PASS
75	95	30	80	48.0	59757	PASS
95	95	100	100	100.0	124565	PASS
96	95	5	9	6.7	8394	PASS
173	174	0.00	2	1.0	1082	PASS
174	95	50	120	87.2	108645	PASS
175	174	5	9	7.3	7948	PASS
176	174	95	101	97.4	105869	PASS
177	176	5	9	6.4	6749	PASS

m/z	Abundance
36.00	821.0
37.10	4510.0
38.00	4253.0
39.10	2072.0
40.10	111.0
44.00	870.0
44.90	828.0
46.00	174.0
47.10	1339.0
48.00	578.0
49.10	4486.0
50.10	20496.0
51.10	6446.0
52.00	360.0
55.00	259.0
56.00	1798.0
57.10	2715.0
60.00	909.0
61.10	5513.0
62.10	5753.0
63.10	4591.0
64.20	708.0
65.10	953.0
67.00	412.0
68.10	11684.0
69.10	12071.0
70.10	601.0
72.10	605.0
73.10	5763.0
74.10	19936.0
75.10	60936.0
76.10	5682.0
77.00	660.0
78.10	483.0
79.00	3400.0
80.00	978.0
81.00	3657.0
82.00	751.0
86.00	197.0
87.00	5723.0
88.00	5252.0
90.90	304.0
92.10	3371.0
93.10	5197.0
94.10	15466.0
95.10	127592.0
96.10	9004.0
97.00	246.0
104.00	428.0
105.00	183.0
105.90	531.0
107.10	113.0
114.90	194.0
116.00	515.0
117.00	705.0
118.00	707.0
119.00	783.0
128.10	381.0
129.10	221.0
130.00	505.0
130.90	262.0
135.00	235.0
136.90	198.0
141.00	1089.0
141.90	107.0
143.10	1711.0
144.20	125.0
145.10	468.0
146.10	201.0
148.00	267.0
153.00	102.0
155.00	344.0
157.00	300.0
159.10	141.0
161.10	128.0
171.10	131.0
172.10	1388.0
173.20	1586.0
174.10	109544.0
175.10	8040.0
176.10	105192.0
177.10	6843.0
178.00	191.0

Instrument :
MSVOA_U
ClientSampleId :
BFB36

m/z	Abundance
36.00	876.0
37.10	4495.0
38.10	4094.0
39.10	2133.0
40.10	118.0
44.00	731.0
44.90	784.0
46.00	183.0
47.10	1326.0
48.10	573.0
49.10	4338.0
50.10	20120.0
51.10	6626.0
52.10	284.0
55.00	350.0
56.00	1760.0
57.10	2656.0
57.90	103.0
60.10	979.0
61.10	5609.0
62.10	5774.0
63.10	4318.0
64.10	662.0
65.10	799.0
67.00	442.0
68.10	11812.0
69.10	11989.0
70.00	578.0
72.10	616.0
73.10	5551.0
74.10	19840.0
75.10	61512.0
76.10	5488.0
77.10	739.0
78.10	494.0
79.00	3398.0
80.00	913.0
81.00	3756.0
82.00	664.0
86.00	225.0
87.00	5814.0
88.00	5558.0
90.90	336.0
92.10	3421.0
93.10	5236.0
94.10	15831.0
95.10	128232.0
96.10	8584.0
97.00	230.0
104.00	477.0
105.00	185.0
105.90	497.0
106.90	123.0
112.90	100.0
114.90	138.0
116.00	447.0
117.00	723.0
118.00	647.0
119.00	698.0
128.00	405.0
129.10	212.0
130.00	465.0
130.90	202.0
134.90	236.0
136.90	168.0
139.90	100.0
141.10	1131.0
143.10	1663.0
144.30	100.0
145.10	460.0
146.10	193.0
148.00	260.0
155.00	396.0
157.00	306.0
159.00	125.0
161.10	126.0
171.00	147.0
172.10	1554.0
173.20	1662.0
174.10	111640.0
175.10	8097.0
176.10	109120.0
177.10	7033.0
178.10	161.0

Instrument :
MSVOA_U
ClientSampleId :
BFB36

m/z	Abundance
36.00	883.0
37.10	3937.0
38.10	3573.0
39.10	2030.0
44.00	668.0
44.90	685.0
46.00	122.0
47.00	1224.0
48.10	474.0
49.10	3836.0
50.10	18400.0
51.10	6141.0
52.10	203.0
55.00	350.0
56.00	1531.0
57.10	2488.0
57.90	136.0
60.10	969.0
61.10	5045.0
62.10	5089.0
63.10	3843.0
64.10	568.0
65.10	705.0
67.00	410.0
68.10	10801.0
69.10	10662.0
70.10	573.0
72.10	580.0
73.10	4883.0
74.10	18056.0
75.10	56824.0
76.10	4824.0
77.10	702.0
78.10	478.0
79.00	3171.0
80.00	860.0
81.00	3311.0
82.00	551.0
85.90	235.0
87.00	5297.0
88.00	5153.0
91.00	304.0
92.10	3020.0
93.10	4973.0
94.10	14643.0
95.10	117872.0
96.10	7594.0
97.00	189.0
104.00	468.0
105.00	143.0
105.90	422.0
106.90	153.0
116.00	340.0
117.00	689.0
118.00	504.0
119.00	561.0
128.00	428.0
129.10	203.0
130.00	398.0
130.90	128.0
134.90	241.0
136.90	155.0
139.90	144.0
141.10	1170.0
143.10	1423.0
145.10	431.0
146.00	164.0
148.00	226.0
150.10	102.0
155.00	398.0
156.90	241.0
159.00	115.0
161.10	113.0
171.00	108.0
172.10	1534.0
174.10	104752.0
175.10	7708.0
176.10	103296.0
177.10	6373.0
178.10	117.0

Instrument :
MSVOA_U
ClientSampleId :
BFB36

m/z	Abundance
44.00	162.0

Instrument :
MSVOA_U
ClientSampleId :
BFB36