

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU010522\
 Data File : VU046666.D
 Acq On : 05 Jan 2022 16:51
 Operator : SY/MD
 Sample : N1009-04
 Misc : 6.73g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGLG0

Quant Time: Jan 06 03:03:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM010522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jan 06 02:54:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	241155	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	243422	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.816	152	122231	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	79568	37.811	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	75.620%
7) Chloroethane-d5	1.919	69	57034	36.483	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	72.960%
11) 1,1-Dichloroethene-d2	2.565	63	106447	29.741	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.480%#
21) 2-Butanone-d5	4.649	46	117026	75.982	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	75.980%
24) Chloroform-d	5.067	84	136589	36.819	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	73.640%
26) 1,2-Dichloroethane-d4	5.703	65	93236	39.629	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	79.260%
32) Benzene-d6	5.729	84	287411	37.834	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	75.660%
36) 1,2-Dichloropropane-d6	6.691	67	87014	37.541	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	75.080%
41) Toluene-d8	7.899	98	258581	38.076	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	76.160%#
43) trans-1,3-Dichloroprop...	8.182	79	44611	39.950	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.900%
47) 2-Hexanone-d5	8.655	63	100053	81.253	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	81.250%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	140680	39.189	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	78.380%
66) 1,2-Dichlorobenzene-d4	12.198	152	107392	42.460	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.920%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU010522\
Data File : VU046666.D
Acq On : 05 Jan 2022 16:51
Operator : SY/MD
Sample : N1009-04
Misc : 6.73g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGLG0

Quant Time: Jan 06 03:03:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM010522WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jan 06 02:54:34 2022
Response via : Initial Calibration

