

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU010522\
 Data File : VU046668.D
 Acq On : 05 Jan 2022 17:37
 Operator : SY/MD
 Sample : N1009-06
 Misc : 5.62g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGLG2

Quant Time: Jan 06 03:04:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM010522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jan 06 02:54:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	282404	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	270664	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	137428	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	125315	50.852	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	101.700%
7) Chloroethane-d5	1.919	69	94842	51.806	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	103.620%
11) 1,1-Dichloroethene-d2	2.572	63	166952	39.833	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.660%
21) 2-Butanone-d5	4.639	46	197474	109.488	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	109.490%
24) Chloroform-d	5.067	84	223498	51.447	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.900%
26) 1,2-Dichloroethane-d4	5.703	65	149556	54.282	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.560%
32) Benzene-d6	5.729	84	461341	54.617	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	109.240%
36) 1,2-Dichloropropane-d6	6.694	67	141918	55.065	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	110.140%
41) Toluene-d8	7.899	98	414233	54.856	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.720%
43) trans-1,3-Dichloroprop...	8.179	79	70134	56.484	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	112.960%
47) 2-Hexanone-d5	8.639	63	168176	122.830	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	122.830%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	221074	55.385	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	110.780%
66) 1,2-Dichlorobenzene-d4	12.195	152	168732	59.335	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	118.680%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU010522\
Data File : VU046668.D
Acq On : 05 Jan 2022 17:37
Operator : SY/MD
Sample : N1009-06
Misc : 5.62g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGLG2

Quant Time: Jan 06 03:04:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM010522WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jan 06 02:54:34 2022
Response via : Initial Calibration

