

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU010522\
 Data File : VU046669.D
 Acq On : 05 Jan 2022 17:59
 Operator : SY/MD
 Sample : N1009-07ME
 Misc : 5.71g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGLG3ME

Quant Time: Jan 06 03:04:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM010522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jan 06 02:54:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	219174	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	215320	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.816	152	109149	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	108989	56.986	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	113.980%
7) Chloroethane-d5	1.919	69	81636	57.457	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	114.920%
11) 1,1-Dichloroethene-d2	2.568	63	148290	45.587	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.180%
21) 2-Butanone-d5	4.645	46	168041	120.047	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	120.050%
24) Chloroform-d	5.067	84	195144	57.879	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	115.760%
26) 1,2-Dichloroethane-d4	5.703	65	132437	61.936	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	123.880%
32) Benzene-d6	5.729	84	402325	59.873	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	119.740%
36) 1,2-Dichloropropane-d6	6.694	67	121398	59.210	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	118.420%
41) Toluene-d8	7.899	98	365716	60.880	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	121.760%#
43) trans-1,3-Dichloroprop...	8.179	79	62011	62.779	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	125.560%#
47) 2-Hexanone-d5	8.645	63	147156	135.103	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	135.100%#
56) 1,1,2,2-Tetrachloroeth...	10.761	84	195697	61.629	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	123.260%#
66) 1,2-Dichlorobenzene-d4	12.195	152	150457	66.617	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	133.240%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU010522\
Data File : VU046669.D
Acq On : 05 Jan 2022 17:59
Operator : SY/MD
Sample : N1009-07ME
Misc : 5.71g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGLG3ME

Quant Time: Jan 06 03:04:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM010522WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jan 06 02:54:34 2022
Response via : Initial Calibration

